

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
 Data File : BE099048.D
 Acq On : 19 Mar 2019 23:37
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 20 05:42:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.341	14.8	101	0.00
3	n-Nitrosodimethylamine	0.400	0.184	54.0#	50	0.05
4 S	2-Fluorophenol	0.400	0.312	22.0	84	0.03
5 S	Phenol-d6	0.400	0.389	2.8	108	0.01
6	bis(2-Chloroethyl)ether	0.400	0.336	16.0	97	0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	128	0.00
8 S	Nitrobenzene-d5	0.400	0.314	21.5	111	0.01
9	Naphthalene	0.400	0.389	2.8	127	-0.01
10	Hexachlorobutadiene	0.400	0.374	6.5	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	127	0.00
12	2-Methylnaphthalene	0.400	0.369	7.8	121	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	131	0.00
14 S	2,4,6-Tribromophenol	0.400	0.249	37.8#	82	0.00
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	129	-0.01
16	Acenaphthylene	0.400	0.390	2.5	134	-0.01
17	Acenaphthene	0.400	0.397	0.8	130	-0.01
18	Fluorene	0.400	0.375	6.3	126	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.400	0.448	-12.0	132	0.00
21	Hexachlorobenzene	0.400	0.398	0.5	121	-0.01
22	Pentachlorophenol	0.400	0.380	5.0	76	0.01
23	Phenanthrene	0.400	0.379	5.3	112	-0.01
24	Anthracene	0.400	0.328	18.0	100	0.00
25 SURR	Fluoranthene-d10	0.400	0.338	15.5	101	0.00
26	Fluoranthene	0.400	0.338	15.5	101	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	96	-0.01
28	Pvrene	0.400	0.419	-4.7	101	0.00
29 S	Terphenyl-d14	0.400	0.381	4.8	92	0.00
30	Benzo(a)anthracene	0.400	0.325	18.8	80	0.00
31	Chrysene	0.400	0.410	-2.5	101	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.337	15.8	88	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.315	21.3	77	0.00
34 I	Perylene-d12	0.400	0.400	0.0	98	-0.02
35	Benzo(b)fluoranthene	0.400	0.363	9.3	92	0.00
36	Benzo(k)fluoranthene	0.400	0.392	2.0	101	-0.01
37 C	Benzo(a)pyrene	0.400	0.376	6.0	96	0.00
38	Dibenzo(a,h)anthracene	0.400	0.247	38.3#	62	0.00
39	Benzo(a,h,i)perylene	0.400	0.309	22.8	78	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
Data File : BE099048.D
Acq On : 19 Mar 2019 23:37
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 05:42:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				