

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
 Data File : BE099048.D
 Acq On : 19 Mar 2019 23:37
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 20 05:42:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.789	0.686	13.1	101	0.00
3	n-Nitrosodimethylamine	1.006	0.462	54.1#	50	0.05
4 S	2-Fluorophenol	1.196	0.933	22.0	84	0.03
5 S	Phenol-d6	1.689	1.644	2.7	108	0.01
6	bis(2-Chloroethyl)ether	1.304	1.094	16.1	97	0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
8 S	Nitrobenzene-d5	0.419	0.329	21.5	111	0.01
9	Naphthalene	4.581	4.458	2.7	127	-0.01
10	Hexachlorobutadiene	0.303	0.283	6.6	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.720	2.6	127	0.00
12	2-Methylnaphthalene	0.743	0.686	7.7	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
14 S	2,4,6-Tribromophenol	0.202	0.126	37.6#	82	0.00
15 S	2-Fluorobiphenyl	1.656	1.617	2.4	129	-0.01
16	Acenaphthylene	2.057	2.008	2.4	134	-0.01
17	Acenaphthene	1.220	1.210	0.8	130	-0.01
18	Fluorene	1.760	1.650	6.3	126	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.217	0.243	-12.0	132	0.00
21	Hexachlorobenzene	0.258	0.256	0.8	121	-0.01
22	Pentachlorophenol	0.052	0.025	51.9#	76	0.01
23	Phenanthrene	1.137	1.077	5.3	112	-0.01
24	Anthracene	0.972	0.796	18.1	100	0.00
25 SURR	Fluoranthene-d10	6.385	5.387	15.6	101	0.00
26	Fluoranthene	1.605	1.357	15.5	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	-0.01
28	Pvrene	1.224	1.283	-4.8	101	0.00
29 S	Terphenyl-d14	0.972	0.925	4.8	92	0.00
30	Benzo(a)anthracene	1.248	1.015	18.7	80	0.00
31	Chrysene	1.336	1.369	-2.5	101	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.235	15.9	88	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.109	21.3	77	0.00
34 I	Perylene-d12	1.000	1.000	0.0	98	-0.02
35	Benzo(b)fluoranthene	1.315	1.193	9.3	92	0.00
36	Benzo(k)fluoranthene	1.377	1.349	2.0	101	-0.01
37 C	Benzo(a)pyrene	1.261	1.187	5.9	96	0.00
38	Dibenzo(a,h)anthracene	1.193	0.737	38.2#	62	0.00
39	Benzo(a,h,i)perylene	1.248	0.964	22.8	78	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
Data File : BE099048.D
Acq On : 19 Mar 2019 23:37
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 05:42:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			