

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
 Data File : BE099065.D
 Acq On : 20 Mar 2019 10:10
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 20 10:45:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.789	0.656	16.9	111	0.00
3	n-Nitrosodimethylamine	1.006	0.827	17.8	104	0.00
4 S	2-Fluorophenol	1.196	1.204	-0.7	125	0.00
5 S	Phenol-d6	1.689	1.700	-0.7	130	0.00
6	bis(2-Chloroethyl)ether	1.304	1.124	13.8	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	133	-0.01
8 S	Nitrobenzene-d5	0.419	0.395	5.7	139	-0.01
9	Naphthalene	4.581	4.590	-0.2	136	-0.01
10	Hexachlorobutadiene	0.303	0.286	5.6	129	-0.01
11 SURR	2-Methylnaphthalene-d10	0.739	0.736	0.4	135	-0.01
12	2-Methylnaphthalene	0.743	0.744	-0.1	137	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
14 S	2,4,6-Tribromophenol	0.202	0.178	11.9	125	-0.01
15 S	2-Fluorobiphenyl	1.656	1.571	5.1	135	-0.01
16	Acenaphthylene	2.057	1.982	3.6	142	-0.01
17	Acenaphthene	1.220	1.158	5.1	134	-0.01
18	Fluorene	1.760	1.647	6.4	135	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	119	-0.01
20	4-Bromophenyl-phenylether	0.217	0.227	-4.6	127	0.00
21	Hexachlorobenzene	0.258	0.250	3.1	122	-0.01
22	Pentachlorophenol	0.052	0.071	-36.5#	221#	-0.01
23	Phenanthrene	1.137	1.095	3.7	118	-0.01
24	Anthracene	0.972	0.936	3.7	122	-0.01
25 SURR	Fluoranthene-d10	6.385	5.663	11.3	110	-0.01
26	Fluoranthene	1.605	1.398	12.9	108	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	104	-0.01
28	Pvrene	1.224	1.284	-4.9	109	-0.01
29 S	Terphenyl-d14	0.972	0.984	-1.2	106	-0.01
30	Benzo(a)anthracene	1.248	1.236	1.0	105	-0.01
31	Chrysene	1.336	1.287	3.7	102	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	3.067	-15.4	131	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.267	10.1	95	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	105	-0.02
35	Benzo(b)fluoranthene	1.315	1.301	1.1	107	-0.01
36	Benzo(k)fluoranthene	1.377	1.301	5.5	104	-0.02
37 C	Benzo(a)pyrene	1.261	1.240	1.7	107	-0.02
38	Dibenzo(a,h)anthracene	1.193	0.919	23.0	82	-0.02
39	Benzo(a,h,i)perylene	1.248	1.009	19.2	87	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031919\
Data File : BE099065.D
Acq On : 20 Mar 2019 10:10
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 10:45:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			