

Data Path : Z:\HPCHEM1\BNA_E\Data\be032116\
 Data File : BE091495.D
 Acq On : 21 Mar 2016 9:51
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 10:24:01 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 11:54:28 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	175#	-0.02
2	1,4-Dioxane	0.624	0.544	12.8	160#	0.00
3	n-Nitrosodimethylamine	0.887	0.831	6.3	189#	-0.02
4 S	2-Fluorophenol	1.149	0.972	15.4	165#	0.00
5 S	Phenol-d6	1.500	1.301	13.3	174#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	171#	-0.02
7 S	Nitrobenzene-d5	0.224	0.195	12.9	177#	-0.01
8	Naphthalene	1.221	1.179	3.4	177#	0.00
9	2-Methylnaphthalene	0.785	0.711	9.4	169#	-0.01
10 I	Acenaphthene-d10	1.000	1.000	0.0	163#	-0.01
11 S	2,4,6-Tribromophenol	0.142	0.137	3.5	194#	-0.01
12 S	2-Fluorobiphenyl	1.258	1.186	5.7	167#	-0.01
13	Acenaphthylene	2.134	1.951	8.6	182#	-0.01
14	Acenaphthene	1.540	1.448	6.0	160#	0.00
15	Fluorene	1.949	1.792	8.1	161#	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	157#	-0.01
17	Hexachlorobenzene	0.218	0.198	9.2	158#	0.00
18	Pentachlorophenol	0.086	0.090	-4.7	203#	-0.02
19	Phenanthrene	1.427	1.350	5.4	160#	-0.01
20	Anthracene	1.252	1.123	10.3	162#	-0.01
21	Fluoranthene	1.638	1.342	18.1	144	-0.02
22 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
23	Pyrene	1.244	1.465	-17.8	151#	-0.02
24 S	Terphenyl-d14	0.601	0.733	-22.0	151#	-0.02
25	Benzo(a)anthracene	1.376	1.465	-6.5	139	0.00
26	Chrysene	1.406	1.853	-31.8#	175#	-0.02
27	Indeno(1,2,3-cd)pyrene	1.434	1.674	-16.7	156#	-0.05
28 I	Perylene-d12	1.000	1.000	0.0	157#	-0.02
29	Benzo(b)fluoranthene	1.580	1.311	17.0	145	-0.02
30	Benzo(k)fluoranthene	1.615	1.401	13.3	148	-0.03
31 C	Benzo(a)pyrene	1.392	1.189	14.6	153#	-0.02
32	Dibenzo(a,h)anthracene	1.371	1.212	11.6	157#	-0.04
33	Benzo(g,h,i)perylene	1.426	1.232	13.6	150	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0