

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032216\
 Data File : BE091505.D
 Acq On : 22 Mar 2016 11:34
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 02:10:31 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 14:14:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	68	0.00
2	1,4-Dioxane	0.400	0.467	-16.8	84	0.00
3	n-Nitrosodimethylamine	0.400	0.355	11.3	70	0.00
4 S	2-Fluorophenol	0.400	0.370	7.5	71	0.00
5 S	Phenol-d6	0.400	0.365	8.8	72	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	64	0.00
7 S	Nitrobenzene-d5	0.400	0.522	-30.5#	74	0.00
8	Naphthalene	0.400	0.449	-12.2	78	0.00
9	2-Methylnaphthalene	0.400	0.395	1.3	70	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	60	0.00
11 S	2,4,6-Tribromophenol	0.400	0.621	-55.2#	92	0.00
12 S	2-Fluorobiphenyl	0.400	0.424	-6.0	69	0.00
13	Acenaphthylene	0.400	0.436	-9.0	67	0.00
14	Acenaphthene	0.400	0.420	-5.0	65	0.00
15	Fluorene	0.400	0.421	-5.2	67	0.00
16 I	Phenanthrene-d10	5.000	5.000	0.0	60	0.00
17	Hexachlorobenzene	0.400	0.423	-5.7	70	0.00
18	Pentachlorophenol	0.400	0.454	-13.5	64	0.00
19	Phenanthrene	0.400	0.433	-8.2	70	0.00
20	Anthracene	0.400	0.389	2.8	67	0.00
21	Fluoranthene	0.400	0.358	10.5	60	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	52	0.00
23	Pyrene	0.400	0.440	-10.0	60	0.00
24 S	Terphenyl-d14	0.400	0.442	-10.5	59	0.00
25	Benzo(a)anthracene	0.400	0.414	-3.5	58	0.00
26	Chrysene	0.400	0.608	-52.0#	79	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.412	-3.0	59	0.00
28 I	Perylene-d12	5.000	5.000	0.0	56	0.00
29	Benzo(b)fluoranthene	0.400	0.393	1.8	61	0.00
30	Benzo(k)fluoranthene	0.400	0.390	2.5	59	0.00
31 C	Benzo(a)pyrene	0.400	0.351	12.3	56	0.00
32	Dibenzo(a,h)anthracene	0.400	0.374	6.5	59	0.00
33	Benzo(g,h,i)perylene	0.400	0.392	2.0	60	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0