

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032216\
 Data File : BE091505.D
 Acq On : 22 Mar 2016 11:34
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 02:10:31 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 18 14:14:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.624	0.729	-16.8	84	0.00
3	n-Nitrosodimethylamine	0.887	0.788	11.2	70	0.00
4 S	2-Fluorophenol	1.149	1.064	7.4	71	0.00
5 S	Phenol-d6	1.500	1.369	8.7	72	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	64	0.00
7 S	Nitrobenzene-d5	0.224	0.216	3.6	74	0.00
8	Naphthalene	1.221	1.371	-12.3	78	0.00
9	2-Methylnaphthalene	0.785	0.775	1.3	70	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	60	0.00
11 S	2,4,6-Tribromophenol	0.142	0.177	-24.6	92	0.00
12 S	2-Fluorobiphenyl	1.258	1.335	-6.1	69	0.00
13	Acenaphthylene	2.134	1.962	8.1	67	0.00
14	Acenaphthene	1.540	1.618	-5.1	65	0.00
15	Fluorene	1.949	2.052	-5.3	67	0.00
16 I	Phenanthrene-d10	1.000	1.000	0.0	60	0.00
17	Hexachlorobenzene	0.218	0.230	-5.5	70	0.00
18	Pentachlorophenol	0.086	0.075	12.8	64	0.00
19	Phenanthrene	1.427	1.544	-8.2	70	0.00
20	Anthracene	1.252	1.218	2.7	67	0.00
21	Fluoranthene	1.638	1.465	10.6	60	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	52	0.00
23	Pyrene	1.244	1.368	-10.0	60	0.00
24 S	Terphenyl-d14	0.601	0.664	-10.5	59	0.00
25	Benzo(a)anthracene	1.376	1.425	-3.6	58	0.00
26	Chrysene	1.406	1.956	-39.1#	79	0.00
27	Indeno(1,2,3-cd)pyrene	1.434	1.479	-3.1	59	0.00
28 I	Perylene-d12	1.000	1.000	0.0	56	0.00
29	Benzo(b)fluoranthene	1.580	1.553	1.7	61	0.00
30	Benzo(k)fluoranthene	1.615	1.575	2.5	59	0.00
31 C	Benzo(a)pyrene	1.392	1.222	12.2	56	0.00
32	Dibenzo(a,h)anthracene	1.371	1.284	6.3	59	0.00
33	Benzo(g,h,i)perylene	1.426	1.397	2.0	60	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0