

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032216\
 Data File : BE091508.D
 Acq On : 22 Mar 2016 20:19
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 02:25:05 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 23 02:10:52 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	65	0.00
2	1,4-Dioxane	0.400	0.482	-20.5	83	0.00
3	n-Nitrosodimethylamine	0.400	0.409	-2.2	77	0.00
4 S	2-Fluorophenol	0.400	0.415	-3.7	76	0.00
5 S	Phenol-d6	0.400	0.409	-2.2	77	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	62	0.00
7 S	Nitrobenzene-d5	0.400	0.560	-40.0#	81	0.00
8	Naphthalene	0.400	0.459	-14.8	76	0.00
9	2-Methylnaphthalene	0.400	0.409	-2.2	70	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	57	-0.01
11 S	2,4,6-Tribromophenol	0.400	0.658	-64.5#	97	0.00
12 S	2-Fluorobiphenyl	0.400	0.443	-10.7	69	0.00
13	Acenaphthylene	0.400	0.454	-13.5	68	0.00
14	Acenaphthene	0.400	0.460	-15.0	69	0.00
15	Fluorene	0.400	0.438	-9.5	67	0.00
16 I	Phenanthrene-d10	5.000	5.000	0.0	58	0.00
17	Hexachlorobenzene	0.400	0.447	-11.7	72	0.00
18	Pentachlorophenol	0.400	0.445	-11.2	60	0.00
19	Phenanthrene	0.400	0.461	-15.3	72	0.00
20	Anthracene	0.400	0.413	-3.2	69	0.00
21	Fluoranthene	0.400	0.396	1.0	64	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	49	0.00
23	Pyrene	0.400	0.470	-17.5	61	0.00
24 S	Terphenyl-d14	0.400	0.480	-20.0	60	0.00
25	Benzo(a)anthracene	0.400	0.463	-15.8	61	0.00
26	Chrysene	0.400	0.638	-59.5#	79	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.426	-6.5	58	0.00
28 I	Perylene-d12	5.000	5.000	0.0	53	-0.01
29	Benzo(b)fluoranthene	0.400	0.417	-4.2	62	0.00
30	Benzo(k)fluoranthene	0.400	0.397	0.8	57	0.00
31 C	Benzo(a)pyrene	0.400	0.365	8.8	55	-0.01
32	Dibenzo(a,h)anthracene	0.400	0.391	2.3	59	0.00
33	Benzo(g,h,i)perylene	0.400	0.406	-1.5	59	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0