

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032216\
 Data File : BE091508.D
 Acq On : 22 Mar 2016 20:19
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 02:25:05 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 23 02:10:52 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
2	1,4-Dioxane	0.624	0.752	-20.5	83	0.00
3	n-Nitrosodimethylamine	0.887	0.908	-2.4	77	0.00
4 S	2-Fluorophenol	1.149	1.192	-3.7	76	0.00
5 S	Phenol-d6	1.500	1.532	-2.1	77	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	62	0.00
7 S	Nitrobenzene-d5	0.224	0.243	-8.5	81	0.00
8	Naphthalene	1.221	1.401	-14.7	76	0.00
9	2-Methylnaphthalene	0.785	0.804	-2.4	70	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	57	-0.01
11 S	2,4,6-Tribromophenol	0.142	0.194	-36.6#	97	0.00
12 S	2-Fluorobiphenyl	1.258	1.393	-10.7	69	0.00
13	Acenaphthylene	2.134	2.081	2.5	68	0.00
14	Acenaphthene	1.540	1.772	-15.1	69	0.00
15	Fluorene	1.949	2.133	-9.4	67	0.00
16 I	Phenanthrene-d10	1.000	1.000	0.0	58	0.00
17	Hexachlorobenzene	0.218	0.243	-11.5	72	0.00
18	Pentachlorophenol	0.086	0.072	16.3	60	0.00
19	Phenanthrene	1.427	1.643	-15.1	72	0.00
20	Anthracene	1.252	1.293	-3.3	69	0.00
21	Fluoranthene	1.638	1.621	1.0	64	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	49#	0.00
23	Pyrene	1.244	1.460	-17.4	61	0.00
24 S	Terphenyl-d14	0.601	0.721	-20.0	60	0.00
25	Benzo(a)anthracene	1.376	1.592	-15.7	61	0.00
26	Chrysene	1.406	2.050	-45.8#	79	0.00
27	Indeno(1,2,3-cd)pyrene	1.434	1.526	-6.4	58	0.00
28 I	Perylene-d12	1.000	1.000	0.0	53	-0.01
29	Benzo(b)fluoranthene	1.580	1.649	-4.4	62	0.00
30	Benzo(k)fluoranthene	1.615	1.602	0.8	57	0.00
31 C	Benzo(a)pyrene	1.392	1.270	8.8	55	-0.01
32	Dibenzo(a,h)anthracene	1.371	1.340	2.3	59	0.00
33	Benzo(g,h,i)perylene	1.426	1.448	-1.5	59	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0