


Data Path : Z:\HPCHEM1\BNA_E\Data\BE032216\
Data File : BE091504.D
Acq On : 22 Mar 2016 10:19
Operator : UM/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE031816.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Mar 23 02:10:52 2016



AutoFind: Scans 2555, 2556, 2557; Background Corrected with Scan 2547

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.9	129387	PASS
68	69	0.00	2	1.7	1918	PASS
69	198	0.00	100	50.6	111215	PASS
70	69	0.00	2	0.5	517	PASS
127	198	10	80	55.7	122273	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	219638	PASS
199	198	5	9	6.8	14874	PASS
275	198	10	60	20.9	45898	PASS
365	198	1	100	2.8	6184	PASS
441	442	0.01	24	19.0	22041	PASS
442	198	50	100	52.8	115885	PASS
443	442	15	24	19.1	22177	PASS