

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032318\
 Data File : BE095827.D
 Acq On : 23 Mar 2018 17:37
 Operator : SJ/JU
 Sample : PB107345BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 50

Instrument :
 BNA_E
 ClientSampleId :
 PB107345BL

Quant Time: Mar 26 16:35:17 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	65300	20.00	ng	0.00
7) Naphthalene-d8	11.26	136	233110	20.00	ng	0.00
13) Acenaphthene-d10	15.00	164	125638	20.00	ng	0.00
19) Phenanthrene-d10	17.71	188	281778	20.00	ng	0.00
27) Chrysene-d12	21.80	240	305857	20.00	ng	0.00
34) Perylene-d12	24.43	264	322885	20.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.92	112	493221	114.56	ng	0.00
5) Phenol-d6	7.57	99	679805	114.83	ng	0.00
8) Nitrobenzene-d5	9.61	82	482706	87.64	ng	0.01
11) 2-Methylnaphthalene-d10	12.80	152	24	0.00	ng	0.00
14) 2,4,6-Tribromophenol	16.47	330	181955	125.48	ng	0.00
15) 2-Fluorobiphenyl	13.64	172	840453	79.36	ng	0.00
25) Fluoranthene-d10	19.68	212	31	0.00	ng	0.00
29) Terphenyl-d14	20.24	244	1130147	86.99	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032318\
Data File : BE095827.D
Acq On : 23 Mar 2018 17:37
Operator : SJ/JU
Sample : PB107345BL
Misc :
ALS Vial : 14 Sample Multiplier: 50

Instrument :
BNA_E
ClientSampleId :
PB107345BL

Quant Time: Mar 26 16:35:17 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

