

Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE032415\  
 Data File : BE089816.D  
 Acq On : 24 Mar 2015 15:31  
 Operator : TP/JJ  
 Sample : SSTDICC002  
 Misc :  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 SSTDICC002

Quant Time: Mar 24 18:26:00 2015  
 Quant Method : Z:\HPCHEM1\BNA\_E\METHODS\SIMPAH-BE032415.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Mar 24 18:03:12 2015  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.50  | 152  | 145031   | 5.00 | ng    | 0.00     |
| 4) Naphthalene-d8         | 10.26 | 136  | 612960   | 5.00 | ng    | 0.00     |
| 8) Acenaphthene-d10       | 14.40 | 164  | 318338   | 5.00 | ng    | 0.00     |
| 13) Phenanthrene-d10      | 17.92 | 188  | 529492   | 5.00 | ng    | 0.00     |
| 19) Chrysene-d12          | 24.26 | 240  | 420354   | 5.00 | ng    | 0.00     |
| 25) Perylene-d12          | 27.41 | 264  | 350352   | 5.00 | ng    | 0.00     |

|                             |       |     |        |      |    |      |
|-----------------------------|-------|-----|--------|------|----|------|
| System Monitoring Compounds |       |     |        |      |    |      |
| 5) Nitrobenzene-d5          | 8.68  | 82  | 61908  | 1.60 | ng | 0.00 |
| 9) 2-Fluorobiphenyl         | 12.85 | 172 | 140435 | 1.89 | ng | 0.00 |
| 21) Terphenyl-d14           | 21.86 | 244 | 116250 | 1.87 | ng | 0.00 |


|                            |       |     |        |        |    |      |
|----------------------------|-------|-----|--------|--------|----|------|
| Target Compounds           |       |     |        | Qvalue |    |      |
| 2) 1,4-Dioxane             | 2.49  | 88  | 38097  | 2.04   | ng | 96   |
| 3) n-Nitrosodimethylamine  | 2.91  | 42  | 57510  | 2.08   | ng | # 95 |
| 6) Naphthalene             | 10.31 | 128 | 224752 | 1.85   | ng | 96   |
| 7) 2-Methylnaphthalene     | 11.93 | 142 | 140559 | 2.09   | ng | 98   |
| 10) Acenaphthylene         | 14.04 | 152 | 857197 | 1.88   | ng | 98   |
| 11) Acenaphthene           | 14.46 | 154 | 126527 | 1.81   | ng | 86   |
| 12) Fluorene               | 15.71 | 166 | 154129 | 1.82   | ng | 94   |
| 14) Hexachlorobenzene      | 16.98 | 284 | 31436  | 1.79   | ng | 99   |
| 15) Pentachlorophenol      | 17.50 | 266 | 7384   | 1.70   | ng | 98   |
| 16) Phenanthrene           | 17.97 | 178 | 207207 | 1.78   | ng | 99   |
| 17) Anthracene             | 18.10 | 178 | 192260 | 1.85   | ng | 100  |
| 18) Fluoranthene           | 20.82 | 202 | 190182 | 1.81   | ng | 98   |
| 20) Pyrene                 | 21.34 | 202 | 195264 | 1.83   | ng | 98   |
| 22) Benzo(a)anthracene     | 24.24 | 228 | 143423 | 1.82   | ng | 97   |
| 23) Chrysene               | 24.32 | 228 | 158102 | 1.84   | ng | 99   |
| 24) Indeno(1,2,3-cd)pyrene | 29.38 | 276 | 54677  | 1.02   | ng | # 67 |
| 26) Benzo(b)fluoranthene   | 26.64 | 252 | 102848 | 1.62   | ng | # 98 |
| 27) Benzo(k)fluoranthene   | 26.70 | 252 | 154175 | 2.15   | ng | # 94 |
| 28) Benzo(a)pyrene         | 27.30 | 252 | 98389  | 1.64   | ng | # 88 |
| 29) Dibenzo(a,h)anthracene | 29.44 | 278 | 85121  | 1.62   | ng | # 79 |
| 30) Benzo(g,h,i)perylene   | 29.83 | 276 | 111914 | 1.72   | ng | # 85 |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE032415\  
Data File : BE089816.D  
Acq On : 24 Mar 2015 15:31  
Operator : TP/JJ  
Sample : SSTDICC002  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
SSTDICC002

Quant Time: Mar 24 18:26:00 2015  
Quant Method : Z:\HPCHEM1\BNA\_E\METHODS\SIMPAH-BE032415.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Mar 24 18:03:12 2015  
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.50 min Scan# 1004

Delta R.T. 0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002


Tgt Ion:152 Resp: 145031

Ion Ratio Lower Upper

152 100

150 153.0 125.5 188.3




115 68.4 58.0 87.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 2.04 ng

RT: 2.49 min Scan# 161

Delta R.T. -0.03 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31


Tgt Ion: 88 Resp: 38097

Ion Ratio Lower Upper

88 100

58 63.4 48.6 72.8



43 32.4 28.6 43.0




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0









#3  
n-Nitrosodimethylamine  
Concen: 2.08 ng  
RT: 2.91 min Scan# 223  
Delta R.T. -0.08 min  
Lab File: BE089816.D  
Acq: 24 Mar 2015 15:31




Instrument :  
BNA\_E  
ClientSampleId :  
SSTDICC002


Tgt Ion: 42 Resp: 57510  
Ion Ratio Lower Upper  
42 100  
74 90.9 68.7 103.1  
44 6.0 6.2 9.4#



#4  
Naphthalene-d8  
Concen: 5.00 ng  
RT: 10.26 min Scan# 1868  
Delta R.T. 0.00 min  
Lab File: BE089816.D  
Acq: 24 Mar 2015 15:31

Tgt Ion: 136 Resp: 612960  
Ion Ratio Lower Upper  
136 100  
137 11.0 8.8 13.2  
54 7.8 7.0 10.4





#5

Nitrobenzene-d5

Concen: 1.60 ng

RT: 8.68 min Scan# 1410

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002


Tgt Ion: 82 Resp: 61908

Ion Ratio Lower Upper

82 100

128 28.9 24.6 37.0

54 51.9 43.0 64.6





Abundance Ion 82.00 (81.70 to 82.70): BE089815.D (-1396) (-)

Ion 128.00 (127.70 to 128.70): BE089815.D (-1396) (-)

Ion 54.00 (53.70 to 54.70): BE089815.D (-1396) (-)

Time--> 8.60 8.65 8.70 8.75



#6

Naphthalene

Concen: 1.85 ng

RT: 10.31 min Scan# 1877

Delta R.T. 0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31


Tgt Ion: 128 Resp: 224752

Ion Ratio Lower Upper

128 100

129 11.2 10.0 15.0

127 13.3 12.5 18.7






Abundance Ion 128.00 (127.70 to 128.70): BE089815.D (-1865) (-)

Ion 129.00 (128.70 to 129.70): BE089815.D (-1865) (-)

Ion 127.00 (126.70 to 127.70): BE089815.D (-1865) (-)

Time--> 10.20 10.30 10.40





#7

2-Methylnaphthalene

Concen: 2.09 ng

RT: 11.93 min Scan# 2206

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002


Tgt Ion:142 Resp: 140559

Ion Ratio Lower Upper

142 100

141 88.9 72.8 109.2




115 42.8 32.9 49.3



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#8

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.40 min Scan# 2622

Delta R.T. 0.01 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31


Tgt Ion:164 Resp: 318338

Ion Ratio Lower Upper

164 100

162 104.0 87.6 131.4



160 42.0 36.7 55.1




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#9

2-Fluorobiphenyl

Concen: 1.89 ng

RT: 12.85 min Scan# 2408

Delta R.T. 0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002


Tgt Ion:172 Resp: 140435

Ion Ratio Lower Upper

172 100

171 35.7 30.8 46.2




170 23.5 21.0 31.4



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#10

Acenaphthylene

Concen: 1.88 ng

RT: 14.04 min Scan# 2590

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31


Tgt Ion:152 Resp: 857197

Ion Ratio Lower Upper

152 100

151 5.1 3.9 5.9



153 12.9 11.0 16.4




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E





#11

Acenaphthene

Concen: 1.81 ng

RT: 14.46 min Scan# 2628

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002


Tgt Ion:154 Resp: 126527

Ion Ratio Lower Upper

154 100

153 464.2 401.5 602.3




152 220.8 198.6 298.0



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#12

Fluorene

Concen: 1.82 ng

RT: 15.71 min Scan# 2742

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31


Tgt Ion:166 Resp: 154129

Ion Ratio Lower Upper

166 100

165 99.6 85.1 127.7



167 13.1 12.2 18.4




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#13  
Phenanthrene-d10  
Concen: 5.00 ng  
RT: 17.92 min Scan# 2972  
Delta R.T. -0.00 min  
Lab File: BE089816.D  
Acq: 24 Mar 2015 15:31

Instrument :  
BNA\_E  
ClientSampleId :  
SSTDICC002

Tgt Ion:188 Resp: 529492  
Ion Ratio Lower Upper  
188 100  
94 9.6 8.5 12.7  
80 11.3 10.0 15.0




Abundance Ion 188.00 (187.70 to 188.70): E  
Ion 94.00 (93.70 to 94.70): BE0  
Ion 80.00 (79.70 to 80.70): BE0






#14  
Hexachlorobenzene  
Concen: 1.79 ng  
RT: 16.98 min Scan# 2868  
Delta R.T. -0.00 min  
Lab File: BE089816.D  
Acq: 24 Mar 2015 15:31

Tgt Ion:284 Resp: 31436  
Ion Ratio Lower Upper  
284 100  
142 67.2 53.3 79.9  
249 38.6 31.1 46.7



Abundance Ion 283.80 (283.50 to 284.50): E  
Ion 142.00 (141.70 to 142.70): E  
Ion 249.00 (248.70 to 249.70): E









#15  
 Pentachlorophenol  
 Concen: 1.70 ng  
 RT: 17.50 min Scan# 2926  
 Delta R.T. 0.00 min  
 Lab File: BE089816.D  
 Acq: 24 Mar 2015 15:31

Instrument :  
 BNA\_E  
 ClientSampleId :  
 SSTDICC002

Tgt Ion:266 Resp: 7384  
 Ion Ratio Lower Upper  
 266 100  
 268 67.2 51.6 77.4  
 264 63.1 50.2 75.2




Abundance Ion 265.70 (265.40 to 266.40): E  
 Ion 268.00 (267.70 to 268.70): E  
 Ion 264.00 (263.70 to 264.70): E






#16  
 Phenanthrene  
 Concen: 1.78 ng  
 RT: 17.97 min Scan# 2978  
 Delta R.T. -0.00 min  
 Lab File: BE089816.D  
 Acq: 24 Mar 2015 15:31

Tgt Ion:178 Resp: 207207  
 Ion Ratio Lower Upper  
 178 100  
 176 19.1 15.4 23.0  
 179 15.7 12.2 18.4



Abundance Ion 178.00 (177.70 to 178.70): E  
 Ion 176.00 (175.70 to 176.70): E  
 Ion 179.00 (178.70 to 179.70): E









#17  
 Anthracene  
 Concen: 1.85 ng  
 RT: 18.10 min Scan# 2992  
 Delta R.T. -0.00 min  
 Lab File: BE089816.D  
 Acq: 24 Mar 2015 15:31




Instrument :  
 BNA\_E  
 ClientSampleId :  
 SSTDICC002


Tgt Ion:178 Resp: 192260  
 Ion Ratio Lower Upper  
 178 100  
 176 18.4 14.6 21.8  
 179 15.1 12.1 18.1



#18  
 Fluoranthene  
 Concen: 1.81 ng  
 RT: 20.82 min Scan# 3432  
 Delta R.T. -0.00 min  
 Lab File: BE089816.D  
 Acq: 24 Mar 2015 15:31

Tgt Ion:202 Resp: 190182  
 Ion Ratio Lower Upper  
 202 100  
 101 9.0 8.2 12.2  
 203 17.4 14.2 21.4





#19

Chrysene-d12

Concen: 5.00 ng

RT: 24.26 min Scan# 4128

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002


Tgt Ion:240 Resp: 420354

Ion Ratio Lower Upper

240 100

120 8.6 7.5 11.3




236 25.7 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#20

Pyrene

Concen: 1.83 ng

RT: 21.34 min Scan# 3530

Delta R.T. -0.01 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31


Tgt Ion:202 Resp: 195264

Ion Ratio Lower Upper

202 100

200 21.0 15.9 23.9



203 17.5 13.6 20.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#21

Terphenyl-d14

Concen: 1.87 ng

RT: 21.86 min Scan# 3628

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002


Tgt Ion:244 Resp: 116250

Ion Ratio Lower Upper

244 100

212 7.0 7.3 10.9#




122 9.4 10.5 15.7#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#22

Benzo(a)anthracene

Concen: 1.82 ng

RT: 24.24 min Scan# 4124

Delta R.T. -0.01 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31


Tgt Ion:228 Resp: 143423

Ion Ratio Lower Upper

228 100

226 26.7 19.8 29.8



229 19.7 16.3 24.5




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#23

Chrysene

Concen: 1.84 ng

RT: 24.32 min Scan# 4141

Delta R.T. 0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002





Tgt Ion: 228 Resp: 158102

Ion Ratio Lower Upper

228 100

226 29.4 23.0 34.4

229 19.8 15.8 23.6



#24

Indeno(1,2,3-cd)pyrene

Concen: 1.02 ng

RT: 29.38 min Scan# 5246

Delta R.T. 0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31




Tgt Ion: 276 Resp: 54677


Ion Ratio Lower Upper

276 100

138 15.0 4.3 6.5#

277 24.7 9.0 13.4#





#25

Perylene-d12

Concen: 5.00 ng

RT: 27.41 min Scan# 4817

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002


Tgt Ion:264 Resp: 350352

Ion Ratio Lower Upper




264 100

260 23.7 18.2 27.2

265 21.9 17.3 25.9



Abundance Ion 264.00 (263.70 to 264.70): E  
Ion 260.00 (259.70 to 260.70): E  
Ion 265.00 (264.70 to 265.70): E



#26

Benzo(b)fluoranthene

Concen: 1.62 ng

RT: 26.64 min Scan# 4647

Delta R.T. 0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31


Tgt Ion:252 Resp: 102848

Ion Ratio Lower Upper



252 100


253 21.9 17.8 26.6

125 8.6 8.7 13.1#



Abundance Ion 252.00 (251.70 to 252.70): E  
Ion 253.00 (252.70 to 253.70): E  
Ion 125.00 (124.70 to 125.70): E





#27

Benzo(k)fluoranthene

Concen: 2.15 ng

RT: 26.70 min Scan# 4660

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002





Tgt Ion:252 Resp: 154175

Ion Ratio Lower Upper

252 100

253 22.0 18.6 28.0

125 8.3 10.8 16.2#



#28

Benzo(a)pyrene

Concen: 1.64 ng

RT: 27.30 min Scan# 4791

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31




Tgt Ion:252 Resp: 98389


Ion Ratio Lower Upper

252 100

253 21.2 21.0 31.4

125 9.5 13.3 19.9#





#29

Dibenzo(a,h)anthracene

Concen: 1.62 ng

RT: 29.44 min Scan# 5258

Delta R.T. -0.00 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Instrument :

BNA\_E

ClientSampleId :

SSTDICC002





Tgt Ion: 278 Resp: 85121

Ion Ratio Lower Upper

278 100

139 13.5 20.0 30.0#

279 24.6 28.9 43.3#



#30

Benzo(g,h,i)perylene

Concen: 1.72 ng

RT: 29.83 min Scan# 5343

Delta R.T. 0.01 min

Lab File: BE089816.D

Acq: 24 Mar 2015 15:31

Tgt Ion: 276 Resp: 111914

Ion Ratio Lower Upper

276 100

277 22.9 19.2 28.8

138 15.3 22.9 34.3#

