

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089813.D
 Acq On : 24 Mar 2015 13:28
 Operator : TP/JJ
 Sample : SSTDICC0.5
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.5

Manual Integrations
 APPROVED

mohammad
 3/25/2015 5:43:35 PM

Quant Time: Mar 24 14:31:42 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 24 14:06:53 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	152202	5.00	ng	0.00
4) Naphthalene-d8	10.26	136	645768	5.00	ng	0.00
8) Acenaphthene-d10	14.40	164	350244	5.00	ng	0.01
13) Phenanthrene-d10	17.92	188	638225	5.00	ng	0.00
19) Chrysene-d12	24.26	240	500309	5.00	ng	0.00
25) Perylene-d12	27.41	264	381761	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.68	82	14403	0.39	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	35113	0.39	ng	0.00
21) Terphenyl-d14	21.86	244	34514	0.50	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.50	88	9588	0.49	ng	# 89
3) n-Nitrosodimethylamine	2.95	42	13419	0.77	ng	# 91
6) Naphthalene	10.31	128	60551	0.45	ng	95
7) 2-Methylnaphthalene	11.93	142	32868	0.42	ng	98
10) Acenaphthylene	14.04	152	219552	0.39	ng	98
11) Acenaphthene	14.46	154	35137	0.41	ng	84
12) Fluorene	15.71	166	43106	0.45	ng	93
14) Hexachlorobenzene	16.98	284	9326	0.47	ng	98
15) Pentachlorophenol	17.50	266	5638	0.07	ng	96
16) Phenanthrene	17.97	178	63250	0.42	ng	99
17) Anthracene	18.10	178	54449	0.42	ng	99
18) Fluoranthene	20.82	202	55352	0.47	ng	98
20) Pyrene	21.34	202	58618	0.35	ng	99
22) Benzo(a)anthracene	24.24	228	39744	0.15	ng	98
23) Chrysene	24.32	228	46519	0.10	ng	99
24) Indeno(1,2,3-cd)pyrene	29.38	276	14637m	0.17	ng	
26) Benzo(b)fluoranthene	26.64	252	21304	0.67	ng	98
27) Benzo(k)fluoranthene	26.70	252	33466m	0.41	ng	
28) Benzo(a)pyrene	27.29	252	19452m	0.51	ng	
29) Dibenzo(a,h)anthracene	29.44	278	10051	0.51	ng	# 85
30) Benzo(g,h,i)perylene	29.83	276	18621	0.12	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
Data File : BE089813.D
Acq On : 24 Mar 2015 13:28
Operator : TP/JJ
Sample : SSTDIC0.5
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.5

Manual Integrations
APPROVED

mohammad
3/25/2015 5:43:35 PM

Quant Time: Mar 24 14:31:42 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 24 14:06:53 2015
Response via : Initial Calibration

