

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089814.D
 Acq On : 24 Mar 2015 14:07
 Operator : TP/JJ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.8

Manual Integrations
 APPROVED

mohammad
 3/25/2015 5:43:33 PM

Quant Time: Mar 24 14:59:12 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 24 14:06:53 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	139944	5.00	ng	0.00
4) Naphthalene-d8	10.26	136	594022	5.00	ng	0.00
8) Acenaphthene-d10	14.40	164	311609	5.00	ng	0.01
13) Phenanthrene-d10	17.92	188	522519	5.00	ng	0.00
19) Chrysene-d12	24.26	240	403320	5.00	ng	0.00
25) Perylene-d12	27.41	264	331224	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.68	82	22764	0.67	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	55157	0.69	ng	0.00
21) Terphenyl-d14	21.86	244	46022	0.83	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.49	88	15054	0.84	ng	93
3) n-Nitrosodimethylamine	2.93	42	19660	1.23	ng	# 82
6) Naphthalene	10.31	128	92852	0.75	ng	95
7) 2-Methylnaphthalene	11.93	142	53155	0.73	ng	98
10) Acenaphthylene	14.04	152	337886	0.68	ng	98
11) Acenaphthene	14.46	154	51833	0.67	ng	87
12) Fluorene	15.70	166	62584	0.73	ng	95
14) Hexachlorobenzene	16.98	284	13272	0.81	ng	98
15) Pentachlorophenol	17.50	266	3485	0.05	ng	94
16) Phenanthrene	17.97	178	86362	0.70	ng	100
17) Anthracene	18.10	178	75997	0.71	ng	99
18) Fluoranthene	20.82	202	75916	0.78	ng	98
20) Pyrene	21.34	202	78094	0.58	ng	98
22) Benzo(a)anthracene	24.24	228	54120	0.25	ng	97
23) Chrysene	24.32	228	63325	0.17	ng	99
24) Indeno(1,2,3-cd)pyrene	29.38	276	26769m	0.39	ng	
26) Benzo(b)fluoranthene	26.64	252	33202	1.20	ng	97
27) Benzo(k)fluoranthene	26.70	252	52003m	0.74	ng	
28) Benzo(a)pyrene	27.30	252	32801m	1.00	ng	
29) Dibenzo(a,h)anthracene	29.44	278	19719	1.16	ng	# 78
30) Benzo(g,h,i)perylene	29.83	276	37741	0.29	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
Data File : BE089814.D
Acq On : 24 Mar 2015 14:07
Operator : TP/JJ
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC0.8

Manual Integrations
APPROVED

mohammad
3/25/2015 5:43:33 PM

Quant Time: Mar 24 14:59:12 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 24 14:06:53 2015
Response via : Initial Calibration

