

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089815.D
 Acq On : 24 Mar 2015 14:49
 Operator : TP/JJ
 Sample : SSTDICCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICCC001

Manual Integrations
 APPROVED

mohammad
 3/25/2015 5:43:40 PM

Quant Time: Mar 24 15:33:34 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 24 14:46:21 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	152830	5.00	ng	0.00
4) Naphthalene-d8	10.26	136	633172	5.00	ng	0.00
8) Acenaphthene-d10	14.40	164	320706	5.00	ng	0.01
13) Phenanthrene-d10	17.92	188	535559	5.00	ng	0.00
19) Chrysene-d12	24.26	240	432336	5.00	ng	0.00
25) Perylene-d12	27.41	264	364287	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.68	82	30534	0.91	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	72755	0.97	ng	0.00
21) Terphenyl-d14	21.86	244	61029	0.92	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.49	88	19336	0.86	ng	93
3) n-Nitrosodimethylamine	2.92	42	27791	1.02	ng	# 90
6) Naphthalene	10.31	128	120366	0.92	ng	96
7) 2-Methylnaphthalene	11.93	142	71928	1.10	ng	98
10) Acenaphthylene	14.04	152	439510	0.97	ng	98
11) Acenaphthene	14.46	154	66890	0.92	ng	86
12) Fluorene	15.70	166	81690	0.94	ng	94
14) Hexachlorobenzene	16.98	284	16948	0.91	ng	98
15) Pentachlorophenol	17.50	266	3717	0.03	ng	95
16) Phenanthrene	17.97	178	111118	0.91	ng	100
17) Anthracene	18.10	178	101407	0.98	ng	100
18) Fluoranthene	20.82	202	99090	0.98	ng	98
20) Pyrene	21.34	202	104223	0.89	ng	99
22) Benzo(a)anthracene	24.24	228	74533	0.94	ng	98
23) Chrysene	24.32	228	86448	0.97	ng	98
24) Indeno(1,2,3-cd)pyrene	29.38	276	43574m	1.32	ng	
26) Benzo(b)fluoranthene	26.64	252	48970	1.10	ng	97
27) Benzo(k)fluoranthene	26.69	252	74238m	1.10	ng	
28) Benzo(a)pyrene	27.30	252	47284	1.14	ng	# 89
29) Dibenzo(a,h)anthracene	29.45	278	37091	1.78	ng	# 78
30) Benzo(g,h,i)perylene	29.83	276	55462	1.27	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
Data File : BE089815.D
Acq On : 24 Mar 2015 14:49
Operator : TP/JJ
Sample : SSTDICCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICCC001

Manual Integrations
APPROVED

mohammad
3/25/2015 5:43:40 PM

Quant Time: Mar 24 15:33:34 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 24 14:46:21 2015
Response via : Initial Calibration

