

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089816.D
 Acq On : 24 Mar 2015 15:31
 Operator : TP/JJ
 Sample : SSTDICC002
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC002

Manual Integrations
 APPROVED

mohammad
 3/25/2015 5:43:30 PM

Quant Time: Mar 24 16:15:01 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 24 15:17:51 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	145031	5.00	ng	0.00
4) Naphthalene-d8	10.26	136	612960	5.00	ng	0.00
8) Acenaphthene-d10	14.40	164	318338	5.00	ng	0.00
13) Phenanthrene-d10	17.92	188	529492	5.00	ng	0.00
19) Chrysene-d12	24.26	240	420354	5.00	ng	0.00
25) Perylene-d12	27.41	264	350352	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.68	82	61908	1.94	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	140435	1.90	ng	0.00
21) Terphenyl-d14	21.86	244	116250	1.84	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.49	88	38097	1.83	ng	96
3) n-Nitrosodimethylamine	2.91	42	57510	2.21	ng	# 95
6) Naphthalene	10.31	128	224752	1.80	ng	96
7) 2-Methylnaphthalene	11.93	142	140559	2.17	ng	98
10) Acenaphthylene	14.04	152	857197	1.92	ng	98
11) Acenaphthene	14.46	154	126527	1.79	ng	86
12) Fluorene	15.71	166	154129	1.81	ng	94
14) Hexachlorobenzene	16.98	284	31436	1.74	ng	99
15) Pentachlorophenol	17.50	266	7384	0.06	ng	98
16) Phenanthrene	17.97	178	207207	1.74	ng	99
17) Anthracene	18.10	178	192260	1.89	ng	100
18) Fluoranthene	20.82	202	190182	1.90	ng	98
20) Pyrene	21.34	202	195264	1.76	ng	98
22) Benzo(a)anthracene	24.24	228	143423	1.88	ng	97
23) Chrysene	24.32	228	158102	1.84	ng	99
24) Indeno(1,2,3-cd)pyrene	29.38	276	96890m	2.84	ng	
26) Benzo(b)fluoranthene	26.64	252	102848	2.35	ng	# 98
27) Benzo(k)fluoranthene	26.70	252	143942m	2.17	ng	
28) Benzo(a)pyrene	27.30	252	98389	2.40	ng	# 88
29) Dibenzo(a,h)anthracene	29.44	278	85121	3.67	ng	# 79
30) Benzo(g,h,i)perylene	29.83	276	111914	2.52	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
Data File : BE089816.D
Acq On : 24 Mar 2015 15:31
Operator : TP/JJ
Sample : SSTDICC002
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC002

Manual Integrations
APPROVED

mohammad
3/25/2015 5:43:30 PM

Quant Time: Mar 24 16:15:01 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 24 15:17:51 2015
Response via : Initial Calibration

