

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089819.D
 Acq On : 24 Mar 2015 18:12
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE032415

Manual Integrations
 APPROVED

mohammad
 3/25/2015 5:43:21 PM

Quant Time: Mar 25 16:40:43 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 24 18:03:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	200642	5.00	ng	0.00
4) Naphthalene-d8	10.26	136	823319	5.00	ng	0.00
8) Acenaphthene-d10	14.39	164	370887	5.00	ng	0.00
13) Phenanthrene-d10	17.92	188	631352	5.00	ng	0.00
19) Chrysene-d12	24.26	240	583671	5.00	ng	0.00
25) Perylene-d12	27.42	264	508863	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.68	82	58403	1.20	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	89126	1.03	ng	0.00
21) Terphenyl-d14	21.86	244	80063	0.93	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.55	88	25853	1.00	ng	# 87
3) n-Nitrosodimethylamine	2.96	42	39768	1.04	ng	# 84
6) Naphthalene	10.31	128	159948	0.98	ng	95
7) 2-Methylnaphthalene	11.93	142	92748	1.03	ng	99
10) Acenaphthylene	14.03	152	551234	1.04	ng	99
11) Acenaphthene	14.46	154	81497	1.00	ng	86
12) Fluorene	15.70	166	99531	1.01	ng	93
14) Hexachlorobenzene	16.98	284	20565	0.98	ng	94
15) Pentachlorophenol	17.51	266	3609	0.56	ng	92
16) Phenanthrene	17.97	178	138914	1.00	ng	100
17) Anthracene	18.10	178	125541	1.01	ng	99
18) Fluoranthene	20.82	202	135382	1.08	ng	99
20) Pyrene	21.34	202	138659	0.94	ng	98
22) Benzo(a)anthracene	24.25	228	106783	0.98	ng	97
23) Chrysene	24.32	228	123666	1.04	ng	100
24) Indeno(1,2,3-cd)pyrene	29.38	276	62208m	0.91	ng	
26) Benzo(b)fluoranthene	26.64	252	64833	0.88	ng	99
27) Benzo(k)fluoranthene	26.70	252	116384	1.12	ng	# 94
28) Benzo(a)pyrene	27.30	252	64200	0.90	ng	# 88
29) Dibenzo(a,h)anthracene	29.44	278	45502	0.86	ng	# 78
30) Benzo(g,h,i)perylene	29.83	276	79586	0.97	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
Data File : BE089819.D
Acq On : 24 Mar 2015 18:12
Operator : TP/JJ
Sample : SSTDICV001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE032415

Manual Integrations
APPROVED

mohammad
3/25/2015 5:43:21 PM

Quant Time: Mar 25 16:40:43 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 24 18:03:12 2015
Response via : Initial Calibration

