

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089824.D
 Acq On : 24 Mar 2015 21:35
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001EC

Manual Integrations
 APPROVED

mohammad
 3/25/2015 5:43:13 PM

Quant Time: Mar 25 03:47:07 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 24 18:03:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	190424	5.00	ng	0.00
4) Naphthalene-d8	10.26	136	778832	5.00	ng	0.00
8) Acenaphthene-d10	14.39	164	357030	5.00	ng	0.00
13) Phenanthrene-d10	17.92	188	606179	5.00	ng	0.00
19) Chrysene-d12	24.26	240	558898	5.00	ng	0.00
25) Perylene-d12	27.42	264	495119	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.68	82	54108	1.18	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	86240	1.03	ng	0.00
21) Terphenyl-d14	21.86	244	78347	0.95	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.54	88	24262	0.99	ng	# 87
3) n-Nitrosodimethylamine	2.96	42	37333	1.03	ng	# 88
6) Naphthalene	10.31	128	150893	0.98	ng	95
7) 2-Methylnaphthalene	11.93	142	90270	1.06	ng	98
10) Acenaphthylene	14.03	152	529371	1.04	ng	99
11) Acenaphthene	14.46	154	79143	1.01	ng	85
12) Fluorene	15.70	166	94949	1.00	ng	94
14) Hexachlorobenzene	16.98	284	19389	0.96	ng	97
15) Pentachlorophenol	17.50	266	4059	0.71	ng	96
16) Phenanthrene	17.97	178	133235	1.00	ng	99
17) Anthracene	18.10	178	122390	1.03	ng	100
18) Fluoranthene	20.82	202	130900	1.09	ng	99
20) Pyrene	21.34	202	136074	0.96	ng	99
22) Benzo(a)anthracene	24.24	228	102146	0.98	ng	98
23) Chrysene	24.32	228	119374	1.05	ng	99
24) Indeno(1,2,3-cd)pyrene	29.38	276	59191m	0.91	ng	
26) Benzo(b)fluoranthene	26.64	252	57010	0.82	ng	98
27) Benzo(k)fluoranthene	26.70	252	116585	1.15	ng	# 94
28) Benzo(a)pyrene	27.30	252	61921	0.89	ng	# 89
29) Dibenzo(a,h)anthracene	29.44	278	48096	0.90	ng	# 80
30) Benzo(g,h,i)perylene	29.83	276	78888	0.98	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
Data File : BE089824.D
Acq On : 24 Mar 2015 21:35
Operator : TP/JJ
Sample : SSTDCCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001EC

Manual Integrations
APPROVED

mohammad
3/25/2015 5:43:13 PM

Quant Time: Mar 25 03:47:07 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 24 18:03:12 2015
Response via : Initial Calibration

