

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089826.D
 Acq On : 25 Mar 2015 00:13
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Manual Integrations
 APPROVED

mohammad
 3/25/2015 5:43:10 PM

Quant Time: Mar 25 17:28:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.50	152	196790	5.00	ng	0.00
4) Naphthalene-d8	10.26	136	801683	5.00	ng	0.00
8) Acenaphthene-d10	14.38	164	363764	5.00	ng	0.00
13) Phenanthrene-d10	17.92	188	616422	5.00	ng	0.00
19) Chrysene-d12	24.26	240	558818	5.00	ng	0.00
25) Perylene-d12	27.41	264	472407	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.68	82	56792	1.20	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	89074	1.05	ng	0.00
21) Terphenyl-d14	21.87	244	78508	0.95	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.55	88	26306	1.03	ng	92
3) n-Nitrosodimethylamine	2.97	42	37641	1.01	ng	# 73
6) Naphthalene	10.30	128	156773	0.99	ng	95
7) 2-Methylnaphthalene	11.93	142	93124	1.06	ng	99
10) Acenaphthylene	14.03	152	547011	1.05	ng	98
11) Acenaphthene	14.46	154	80145	1.01	ng	87
12) Fluorene	15.70	166	97278	1.01	ng	93
14) Hexachlorobenzene	16.98	284	20533	1.00	ng	94
15) Pentachlorophenol	17.50	266	4502	0.80	ng	95
16) Phenanthrene	17.97	178	136820	1.01	ng	99
17) Anthracene	18.10	178	125701	1.04	ng	100
18) Fluoranthene	20.82	202	132828	1.09	ng	99
20) Pyrene	21.34	202	137492	0.97	ng	99
22) Benzo(a)anthracene	24.25	228	96517	0.92	ng	97
23) Chrysene	24.32	228	116103	1.02	ng	98
24) Indeno(1,2,3-cd)pyrene	29.38	276	33989m	0.70	ng	
26) Benzo(b)fluoranthene	26.64	252	42475	0.71	ng	100
27) Benzo(k)fluoranthene	26.70	252	98692	1.02	ng	# 94
28) Benzo(a)pyrene	27.30	252	47479	0.78	ng	# 90
29) Dibenzo(a,h)anthracene	29.44	278	26232	0.64	ng	# 80
30) Benzo(g,h,i)perylene	29.83	276	64365	0.88	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
Data File : BE089826.D
Acq On : 25 Mar 2015 00:13
Operator : TP/JJ
Sample : SSTDCCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC001

Manual Integrations
APPROVED

mohammad
3/25/2015 5:43:10 PM

Quant Time: Mar 25 17:28:38 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 25 16:20:31 2015
Response via : Initial Calibration

