

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089819.D
 Acq On : 24 Mar 2015 18:12
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE032415

Quant Time: Mar 25 16:40:43 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 24 18:03:12 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	0.00
2	1,4-Dioxane	0.689	0.644	6.5	134	0.03
3	n-Nitrosodimethylamine	0.951	0.991	-4.2	143	-0.03
4 I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
5 S	Nitrobenzene-d5	0.284	0.355	-25.0#	191#	0.00
6	Naphthalene	0.988	0.971	1.7	133	0.00
7	2-Methylnaphthalene	0.549	0.563	-2.6	129	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
9 S	2-Fluorobiphenyl	1.170	1.202	-2.7	123	0.00
10	Acenaphthylene	7.154	7.431	-3.9	125	-0.01
11	Acenaphthene	1.095	1.099	-0.4	122	0.00
12	Fluorene	1.328	1.342	-1.1	122	-0.01
13 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
14	Hexachlorobenzene	0.166	0.163	1.8	121	0.00
15	Pentachlorophenol	0.053	0.029	45.3#	97	0.00
16	Phenanthrene	1.099	1.100	-0.1	125	0.00
17	Anthracene	0.982	0.994	-1.2	124	0.00
18	Fluoranthene	0.991	1.072	-8.2	137	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	135	0.00
20	Pyrene	1.268	1.188	6.3	133	0.00
21 S	Terphenyl-d14	0.738	0.686	7.0	131	0.00
22	Benzo(a)anthracene	0.935	0.915	2.1	143	0.00
23	Chrysene	1.020	1.059	-3.8	143	0.00
24	Indeno(1,2,3-cd)pyrene	0.578	0.533	7.8	143	0.00
25 I	Perylene-d12	1.000	1.000	0.0	140	0.00
26	Benzo(b)fluoranthene	0.745	0.637	14.5	132	0.00
27	Benzo(k)fluoranthene	1.024	1.144	-11.7	157#	0.00
28 C	Benzo(a)pyrene	0.702	0.631	10.1	136	0.00
29	Dibenzo(a,h)anthracene	0.513	0.447	12.9	123	0.00
30	Benzo(g,h,i)perylene	0.756	0.782	-3.4	143	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0