

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089826.D
 Acq On : 25 Mar 2015 00:13
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 25 17:28:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	129	0.00
2	1,4-Dioxane	1.000	1.035	-3.5	136	0.02
3	n-Nitrosodimethylamine	1.000	1.005	-0.5	135	-0.03
4 I	Naphthalene-d8	5.000	5.000	0.0	127	0.00
5 S	Nitrobenzene-d5	1.000	1.199	-19.9	186	0.00
6	Naphthalene	1.000	0.989	1.1	130	0.00
7	2-Methylnaphthalene	1.000	1.058	-5.8	129	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	113	0.00
9 S	2-Fluorobiphenyl	1.000	1.046	-4.6	122	0.00
10	Acenaphthylene	1.000	1.051	-5.1	124	-0.01
11	Acenaphthene	1.000	1.006	-0.6	120	0.00
12	Fluorene	1.000	1.007	-0.7	119	-0.01
13 I	Phenanthrene-d10	5.000	5.000	0.0	115	0.00
14	Hexachlorobenzene	1.000	1.003	-0.3	121	0.00
15	Pentachlorophenol	1.000	0.799	20.1	121	0.00
16	Phenanthrene	1.000	1.010	-1.0	123	0.00
17	Anthracene	1.000	1.038	-3.8	124	0.00
18	Fluoranthene	1.000	1.087	-8.7	134	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	129	0.00
20	Pyrene	1.000	0.970	3.0	132	0.00
21 S	Terphenyl-d14	1.000	0.952	4.8	129	0.00
22	Benzo(a)anthracene	1.000	0.923	7.7	129	0.00
23	Chrysene	1.000	1.018	-1.8	134	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.702	29.8#	78	0.00
25 I	Perylene-d12	5.000	5.000	0.0	130	0.00
26	Benzo(b)fluoranthene	1.000	0.712	28.8#	87	0.00
27	Benzo(k)fluoranthene	1.000	1.020	-2.0	133	0.00
28 C	Benzo(a)pyrene	1.000	0.775	22.5#	100	0.00
29	Dibenzo(a,h)anthracene	1.000	0.635	36.5#	71	0.00
30	Benzo(g,h,i)perylene	1.000	0.877	12.3	116	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1