

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089826.D
 Acq On : 25 Mar 2015 00:13
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 25 17:28:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.689	0.668	3.0	136	0.02
3	n-Nitrosodimethylamine	0.951	0.956	-0.5	135	-0.03
4 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
5 S	Nitrobenzene-d5	0.284	0.354	-24.6	186#	0.00
6	Naphthalene	0.988	0.978	1.0	130	0.00
7	2-Methylnaphthalene	0.549	0.581	-5.8	129	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
9 S	2-Fluorobiphenyl	1.170	1.224	-4.6	122	0.00
10	Acenaphthylene	7.154	7.519	-5.1	124	-0.01
11	Acenaphthene	1.095	1.102	-0.6	120	0.00
12	Fluorene	1.328	1.337	-0.7	119	-0.01
13 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
14	Hexachlorobenzene	0.166	0.167	-0.6	121	0.00
15	Pentachlorophenol	0.053	0.037	30.2#	121	0.00
16	Phenanthrene	1.099	1.110	-1.0	123	0.00
17	Anthracene	0.982	1.020	-3.9	124	0.00
18	Fluoranthene	0.991	1.077	-8.7	134	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
20	Pyrene	1.268	1.230	3.0	132	0.00
21 S	Terphenyl-d14	0.738	0.702	4.9	129	0.00
22	Benzo(a)anthracene	0.935	0.864	7.6	129	0.00
23	Chrysene	1.020	1.039	-1.9	134	0.00
24	Indeno(1,2,3-cd)pyrene	0.578	0.304	47.4#	78	0.00
25 I	Perylene-d12	1.000	1.000	0.0	130	0.00
26	Benzo(b)fluoranthene	0.745	0.450	39.6#	87	0.00
27	Benzo(k)fluoranthene	1.024	1.045	-2.1	133	0.00
28 C	Benzo(a)pyrene	0.702	0.503	28.3#	100	0.00
29	Dibenzo(a,h)anthracene	0.513	0.278	45.8#	71	0.00
30	Benzo(g,h,i)perylene	0.756	0.681	9.9	116	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1