

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089838.D
 Acq On : 25 Mar 2015 8:58
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 25 17:25:36 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	118	0.00
2	1,4-Dioxane	1.000	-0.008	100.8#	0	-2.52#
3	n-Nitrosodimethylamine	1.000	0.092	90.8#	11	0.04
4 I	Naphthalene-d8	5.000	5.000	0.0	112	0.00
5 S	Nitrobenzene-d5	1.000	0.373	62.7#	20	0.00
6	Naphthalene	1.000	0.112	88.8#	13	0.00
7	2-Methylnaphthalene	1.000	0.112	88.8#	12	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.01
9 S	2-Fluorobiphenyl	1.000	0.134	86.6#	14	0.00
10	Acenaphthylene	1.000	0.109	89.1#	11	0.00
11	Acenaphthene	1.000	0.111	88.9#	12	0.00
12	Fluorene	1.000	0.111	88.9#	12	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	103	0.00
14	Hexachlorobenzene	1.000	0.112	88.8#	12	0.00
15	Pentachlorophenol	1.000	-1.000	200.0#	0	-17.50#
16	Phenanthrene	1.000	0.115	88.5#	13	0.00
17	Anthracene	1.000	0.109	89.1#	12	0.00
18	Fluoranthene	1.000	0.118	88.2#	13	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	118	0.00
20	Pyrene	1.000	0.102	89.8#	13	0.00
21 S	Terphenyl-d14	1.000	0.113	88.7#	14	0.00
22	Benzo(a)anthracene	1.000	0.126	87.4#	16	0.00
23	Chrysene	1.000	0.112	88.8#	14	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.493	50.7#	18	0.00
25 I	Perylene-d12	5.000	5.000	0.0	126	0.00
26	Benzo(b)fluoranthene	1.000	0.379	62.1#	14	0.00
27	Benzo(k)fluoranthene	1.000	0.112	88.8#	14	0.00
28 C	Benzo(a)pyrene	1.000	0.373	62.7#	16	0.00
29	Dibenzo(a,h)anthracene	1.000	0.318	68.2#	11	0.00
30	Benzo(g,h,i)perylene	1.000	0.337	66.3#	15	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1