

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032415\
 Data File : BE089838.D
 Acq On : 25 Mar 2015 8:58
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 25 17:25:36 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2	1,4-Dioxane	0.689	0.000	100.0#	0#	-2.52#
3	n-Nitrosodimethylamine	0.951	0.087	90.9#	11#	0.04
4 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
5 S	Nitrobenzene-d5	0.284	0.043	84.9#	20#	0.00
6	Naphthalene	0.988	0.111	88.8#	13#	0.00
7	2-Methylnaphthalene	0.549	0.061	88.9#	12#	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.01
9 S	2-Fluorobiphenyl	1.170	0.157	86.6#	14#	0.00
10	Acenaphthylene	7.154	0.777	89.1#	11#	0.00
11	Acenaphthene	1.095	0.122	88.9#	12#	0.00
12	Fluorene	1.328	0.147	88.9#	12#	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
14	Hexachlorobenzene	0.166	0.019	88.6#	12#	0.00
15	Pentachlorophenol	0.053	0.000	100.0#	0#	-17.50#
16	Phenanthrene	1.099	0.127	88.4#	13#	0.00
17	Anthracene	0.982	0.107	89.1#	12#	0.00
18	Fluoranthene	0.991	0.117	88.2#	13#	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
20	Pyrene	1.268	0.129	89.8#	13#	0.00
21 S	Terphenyl-d14	0.738	0.083	88.8#	14#	0.00
22	Benzo(a)anthracene	0.935	0.117	87.5#	16#	0.00
23	Chrysene	1.020	0.115	88.7#	14#	0.00
24	Indeno(1,2,3-cd)pyrene	0.578	0.077	86.7#	18#	0.00
25 I	Perylene-d12	1.000	1.000	0.0	126	0.00
26	Benzo(b)fluoranthene	0.745	0.076	89.8#	14#	0.00
27	Benzo(k)fluoranthene	1.024	0.114	88.9#	14#	0.00
28 C	Benzo(a)pyrene	0.702	0.081	88.5#	16#	0.00
29	Dibenzo(a,h)anthracene	0.513	0.043	91.6#	11#	0.00
30	Benzo(g,h,i)perylene	0.756	0.092	87.8#	15#	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1