

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032515\
 Data File : BE089840.D
 Acq On : 25 Mar 2015 16:37
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 25 17:28:01 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	167#	0.00
2	1,4-Dioxane	0.689	0.656	4.8	174#	0.03
3	n-Nitrosodimethylamine	0.951	0.936	1.6	172#	-0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	162#	0.00
5 S	Nitrobenzene-d5	0.284	0.362	-27.5#	244#	0.00
6	Naphthalene	0.988	0.983	0.5	168#	0.00
7	2-Methylnaphthalene	0.549	0.581	-5.8	166#	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	138	0.00
9 S	2-Fluorobiphenyl	1.170	1.273	-8.8	155#	0.00
10	Acenaphthylene	7.154	7.652	-7.0	155#	0.00
11	Acenaphthene	1.095	1.139	-4.0	151#	0.00
12	Fluorene	1.328	1.357	-2.2	148	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	138	0.00
14	Hexachlorobenzene	0.166	0.165	0.6	144	0.00
15	Pentachlorophenol	0.053	0.032	39.6#	126	0.00
16	Phenanthrene	1.099	1.104	-0.5	147	0.00
17	Anthracene	0.982	1.009	-2.7	147	0.00
18	Fluoranthene	0.991	1.061	-7.1	158#	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	151#	0.00
20	Pyrene	1.268	1.237	2.4	155#	0.00
21 S	Terphenyl-d14	0.738	0.731	0.9	157#	0.00
22	Benzo(a)anthracene	0.935	0.926	1.0	163#	0.00
23	Chrysene	1.020	1.076	-5.5	163#	0.00
24	Indeno(1,2,3-cd)pyrene	0.578	0.590	-2.1	177#	0.00
25 I	Perylene-d12	1.000	1.000	0.0	157#	0.00
26	Benzo(b)fluoranthene	0.745	0.630	15.4	147	0.00
27	Benzo(k)fluoranthene	1.024	1.159	-13.2	179#	0.00
28 C	Benzo(a)pyrene	0.702	0.653	7.0	159#	0.00
29	Dibenzo(a,h)anthracene	0.513	0.455	11.3	141	0.00
30	Benzo(g,h,i)perylene	0.756	0.859	-13.6	178#	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0