

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032515\
 Data File : BE089840.D
 Acq On : 25 Mar 2015 16:37
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 25 17:28:01 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	167	0.00
2	1,4-Dioxane	1.000	1.016	-1.6	174	0.03
3	n-Nitrosodimethylamine	1.000	0.984	1.6	172	-0.02
4 I	Naphthalene-d8	5.000	5.000	0.0	162	0.00
5 S	Nitrobenzene-d5	1.000	1.219	-21.9	244	0.00
6	Naphthalene	1.000	0.995	0.5	168	0.00
7	2-Methylnaphthalene	1.000	1.059	-5.9	166	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	138	0.00
9 S	2-Fluorobiphenyl	1.000	1.088	-8.8	155	0.00
10	Acenaphthylene	1.000	1.070	-7.0	155	0.00
11	Acenaphthene	1.000	1.040	-4.0	151	0.00
12	Fluorene	1.000	1.022	-2.2	148	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	138	0.00
14	Hexachlorobenzene	1.000	0.994	0.6	144	0.00
15	Pentachlorophenol	1.000	0.653	34.7#	126	0.00
16	Phenanthrene	1.000	1.005	-0.5	147	0.00
17	Anthracene	1.000	1.027	-2.7	147	0.00
18	Fluoranthene	1.000	1.071	-7.1	158	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	151	0.00
20	Pyrene	1.000	0.975	2.5	155	0.00
21 S	Terphenyl-d14	1.000	0.991	0.9	157	0.00
22	Benzo(a)anthracene	1.000	0.990	1.0	163	0.00
23	Chrysene	1.000	1.054	-5.4	163	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.964	3.6	177	0.00
25 I	Perylene-d12	5.000	5.000	0.0	157	0.00
26	Benzo(b)fluoranthene	1.000	0.873	12.7	147	0.00
27	Benzo(k)fluoranthene	1.000	1.132	-13.2	179	0.00
28 C	Benzo(a)pyrene	1.000	0.919	8.1	159	0.00
29	Dibenzo(a,h)anthracene	1.000	0.870	13.0	141	0.00
30	Benzo(g,h,i)perylene	1.000	1.041	-4.1	178	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0