

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032515\
 Data File : BE089855.D
 Acq On : 26 Mar 2015 2:51
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Mar 26 11:53:53 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2	1,4-Dioxane	0.689	0.589	14.5	111	0.00
3	n-Nitrosodimethylamine	0.951	0.852	10.4	112	-0.05
4 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
5 S	Nitrobenzene-d5	0.284	0.328	-15.5	148	0.00
6	Naphthalene	0.988	0.986	0.2	113	0.00
7	2-Methylnaphthalene	0.549	0.585	-6.6	112	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.01
9 S	2-Fluorobiphenyl	1.170	1.273	-8.8	106	0.00
10	Acenaphthylene	7.154	7.621	-6.5	105	0.00
11	Acenaphthene	1.095	1.113	-1.6	100	0.00
12	Fluorene	1.328	1.337	-0.7	99	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
14	Hexachlorobenzene	0.166	0.167	-0.6	99	0.00
15	Pentachlorophenol	0.053	0.030	43.4#	82	0.00
16	Phenanthrene	1.099	1.106	-0.6	100	0.00
17	Anthracene	0.982	1.034	-5.3	102	0.00
18	Fluoranthene	0.991	1.091	-10.1	111	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
20	Pyrene	1.268	1.183	6.7	109	0.00
21 S	Terphenyl-d14	0.738	0.736	0.3	116	0.00
22	Benzo(a)anthracene	0.935	1.022	-9.3	131	0.00
23	Chrysene	1.020	1.062	-4.1	118	0.00
24	Indeno(1,2,3-cd)pyrene	0.578	0.919	-59.0#	202#	0.00
25 I	Perylene-d12	1.000	1.000	0.0	126	0.00
26	Benzo(b)fluoranthene	0.745	0.902	-21.1	170#	0.00
27	Benzo(k)fluoranthene	1.024	1.140	-11.3	142	0.00
28 C	Benzo(a)pyrene	0.702	0.912	-29.9#	178#	0.00
29	Dibenzo(a,h)anthracene	0.513	0.749	-46.0#	186#	0.00
30	Benzo(g,h,i)perylene	0.756	1.006	-33.1#	167#	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1