

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032515\
 Data File : BE089855.D
 Acq On : 26 Mar 2015 2:51
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001

Quant Time: Mar 26 11:53:53 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	119	0.00
2	1,4-Dioxane	1.000	0.911	8.9	111	0.00
3	n-Nitrosodimethylamine	1.000	0.896	10.4	112	-0.05
4 I	Naphthalene-d8	5.000	5.000	0.0	109	0.00
5 S	Nitrobenzene-d5	1.000	1.130	-13.0	148	0.00
6	Naphthalene	1.000	0.997	0.3	113	0.00
7	2-Methylnaphthalene	1.000	1.065	-6.5	112	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	94	0.01
9 S	2-Fluorobiphenyl	1.000	1.088	-8.8	106	0.00
10	Acenaphthylene	1.000	1.065	-6.5	105	0.00
11	Acenaphthene	1.000	1.016	-1.6	100	0.00
12	Fluorene	1.000	1.006	-0.6	99	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	94	0.00
14	Hexachlorobenzene	1.000	1.008	-0.8	99	0.00
15	Pentachlorophenol	1.000	0.607	39.3#	82	0.00
16	Phenanthrene	1.000	1.006	-0.6	100	0.00
17	Anthracene	1.000	1.053	-5.3	102	0.00
18	Fluoranthene	1.000	1.101	-10.1	111	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	111	0.00
20	Pyrene	1.000	0.933	6.7	109	0.00
21 S	Terphenyl-d14	1.000	0.998	0.2	116	0.00
22	Benzo(a)anthracene	1.000	1.093	-9.3	131	0.00
23	Chrysene	1.000	1.041	-4.1	118	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	1.266	-26.6#	202	0.00
25 I	Perylene-d12	5.000	5.000	0.0	126	0.00
26	Benzo(b)fluoranthene	1.000	1.116	-11.6	170	0.00
27	Benzo(k)fluoranthene	1.000	1.113	-11.3	142	0.00
28 C	Benzo(a)pyrene	1.000	1.166	-16.6	178	0.00
29	Dibenzo(a,h)anthracene	1.000	1.248	-24.8	186	0.00
30	Benzo(g,h,i)perylene	1.000	1.175	-17.5	167	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0