

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032516\
 Data File : BE091530.D
 Acq On : 25 Mar 2016 18:26
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE032516

Quant Time: Mar 25 19:06:13 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 18:30:57 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.669	0.688	-2.8	92	0.00
3	n-Nitrosodimethylamine	0.581	0.592	-1.9	97	0.00
4 S	2-Fluorophenol	0.975	0.960	1.5	99	0.00
5 S	Phenol-d6	1.330	1.259	5.3	99	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
7 S	Nitrobenzene-d5	0.217	0.219	-0.9	111	-0.01
8	Naphthalene	1.007	1.015	-0.8	96	0.00
9	2-Methylnaphthalene	0.629	0.613	2.5	97	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	93	-0.01
11 S	2,4,6-Tribromophenol	0.085	0.084	1.2	107	-0.01
12 S	2-Fluorobiphenyl	1.408	1.420	-0.9	96	-0.01
13	Acenaphthylene	1.752	1.647	6.0	99	0.00
14	Acenaphthene	1.237	1.251	-1.1	95	-0.01
15	Fluorene	1.511	1.442	4.6	94	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
17	Hexachlorobenzene	0.176	0.170	3.4	96	0.00
18	Pentachlorophenol	0.047	0.040	14.9	98	-0.02
19	Phenanthrene	1.145	1.134	1.0	94	0.00
20	Anthracene	0.948	0.881	7.1	98	0.00
21	Fluoranthene	1.071	0.991	7.5	96	-0.02
22 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
23	Pyrene	1.181	1.083	8.3	95	0.00
24 S	Terphenyl-d14	0.822	0.762	7.3	94	0.00
25	Benzo(a)anthracene	1.154	1.057	8.4	95	0.00
26	Chrysene	1.516	1.538	-1.5	95	0.00
27	Indeno(1,2,3-cd)pyrene	1.179	1.066	9.6	96	-0.02
28 I	Perylene-d12	1.000	1.000	0.0	96	-0.01
29	Benzo(b)fluoranthene	1.111	0.994	10.5	97	-0.01
30	Benzo(k)fluoranthene	1.107	1.053	4.9	95	0.00
31 C	Benzo(a)pyrene	0.926	0.789	14.8	96	-0.01
32	Dibenzo(a,h)anthracene	0.984	0.886	10.0	96	-0.02
33	Benzo(g,h,i)perylene	1.029	0.948	7.9	96	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0