

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032516\
 Data File : BE091530.D
 Acq On : 25 Mar 2016 18:26
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE032516

Quant Time: Mar 25 19:06:13 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 18:30:57 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	0.00
2	1,4-Dioxane	0.400	0.414	-3.5	92	0.00
3	n-Nitrosodimethylamine	0.400	0.409	-2.2	97	0.00
4 S	2-Fluorophenol	0.400	0.394	1.5	99	0.00
5 S	Phenol-d6	0.400	0.379	5.3	99	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	94	0.00
7 S	Nitrobenzene-d5	0.400	0.403	-0.8	111	-0.01
8	Naphthalene	0.400	0.403	-0.8	96	0.00
9	2-Methylnaphthalene	0.400	0.389	2.8	97	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	93	-0.01
11 S	2,4,6-Tribromophenol	0.400	0.395	1.3	107	-0.01
12 S	2-Fluorobiphenyl	0.400	0.403	-0.8	96	-0.01
13	Acenaphthylene	0.400	0.376	6.0	99	0.00
14	Acenaphthene	0.400	0.404	-1.0	95	-0.01
15	Fluorene	0.400	0.382	4.5	94	-0.01
16 I	Phenanthrene-d10	5.000	5.000	0.0	95	0.00
17	Hexachlorobenzene	0.400	0.386	3.5	96	0.00
18	Pentachlorophenol	0.400	0.434	-8.5	98	-0.02
19	Phenanthrene	0.400	0.396	1.0	94	0.00
20	Anthracene	0.400	0.372	7.0	98	0.00
21	Fluoranthene	0.400	0.370	7.5	96	-0.02
22 I	Chrysene-d12	5.000	5.000	0.0	94	0.00
23	Pyrene	0.400	0.367	8.3	95	0.00
24 S	Terphenyl-d14	0.400	0.371	7.3	94	0.00
25	Benzo(a)anthracene	0.400	0.366	8.5	95	0.00
26	Chrysene	0.400	0.406	-1.5	95	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.362	9.5	96	-0.02
28 I	Perylene-d12	5.000	5.000	0.0	96	-0.01
29	Benzo(b)fluoranthene	0.400	0.358	10.5	97	-0.01
30	Benzo(k)fluoranthene	0.400	0.381	4.8	95	0.00
31 C	Benzo(a)pyrene	0.400	0.477	-19.2	96	-0.01
32	Dibenzo(a,h)anthracene	0.400	0.360	10.0	96	-0.02
33	Benzo(g,h,i)perylene	0.400	0.368	8.0	96	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0