

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032615\
 Data File : BE089862.D
 Acq On : 26 Mar 2015 18:09
 Operator : TP/JJ
 Sample : SSTDICC0.5
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.5

Quant Time: Mar 27 02:13:10 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 27 01:45:40 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	84710	5.00	ng	0.00
4) Naphthalene-d8	10.23	136	369052	5.00	ng	0.00
8) Acenaphthene-d10	14.36	164	200316	5.00	ng	0.00
13) Phenanthrene-d10	17.89	188	383316	5.00	ng	0.00
19) Chrysene-d12	24.23	240	340531	5.00	ng	0.00
25) Perylene-d12	27.38	264	319822	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.66	82	9726	0.45	ng	0.00
9) 2-Fluorobiphenyl	12.83	172	23993	0.51	ng	0.00
21) Terphenyl-d14	21.83	244	27083	0.54	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.47	88	5121	0.44	ng	89
3) n-Nitrosodimethylamine	2.93	42	6009	0.38	ng	# 60
6) Naphthalene	10.28	128	37312	0.50	ng	97
7) 2-Methylnaphthalene	11.91	142	23419	0.56	ng	99
10) Acenaphthylene	14.01	152	140638	0.48	ng	99
11) Acenaphthene	14.44	154	21333	0.48	ng	88
12) Fluorene	15.68	166	27838	0.51	ng	93
14) Hexachlorobenzene	16.95	284	5664	0.44	ng	98
15) Pentachlorophenol	17.48	266	11089	2.38	ng	95
16) Phenanthrene	17.95	178	41198	0.48	ng	98
17) Anthracene	18.07	178	35386	0.46	ng	99
18) Fluoranthene	20.79	202	39517	0.51	ng	98
20) Pyrene	21.31	202	41962	0.49	ng	99
22) Benzo(a)anthracene	24.21	228	34380	0.53	ng	96
23) Chrysene	24.28	228	35449	0.51	ng	99
24) Indeno(1,2,3-cd)pyrene	29.35	276	31922m	0.76	ng	
26) Benzo(b)fluoranthene	26.60	252	53761	1.04	ng	98
27) Benzo(k)fluoranthene	26.66	252	68953m	1.00	ng	
28) Benzo(a)pyrene	27.26	252	50550m	1.04	ng	
29) Dibenzo(a,h)anthracene	29.41	278	22032	0.63	ng	# 82
30) Benzo(g,h,i)perylene	29.79	276	30287	0.61	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032615\
Data File : BE089862.D
Acq On : 26 Mar 2015 18:09
Operator : TP/JJ
Sample : SSTDIC0.5
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.5

Quant Time: Mar 27 02:13:10 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 27 01:45:40 2015
Response via : Initial Calibration

