

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032615\
 Data File : BE089866.D
 Acq On : 26 Mar 2015 20:53
 Operator : TP/JJ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Quant Time: Mar 27 02:17:26 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 27 01:45:40 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	86522	5.00	ng	0.00
4) Naphthalene-d8	10.23	136	364164	5.00	ng	0.00
8) Acenaphthene-d10	14.36	164	217664	5.00	ng	0.00
13) Phenanthrene-d10	17.89	188	404435	5.00	ng	0.00
19) Chrysene-d12	24.23	240	421154	5.00	ng	0.00
25) Perylene-d12	27.38	264	377578	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.66	82	110326	5.23	ng	0.00
9) 2-Fluorobiphenyl	12.83	172	232544	4.51	ng	0.00
21) Terphenyl-d14	21.83	244	269803	4.32	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.45	88	53982	4.53	ng	92
3) n-Nitrosodimethylamine	2.88	42	71847	4.46	ng	83
6) Naphthalene	10.28	128	350092	4.79	ng	96
7) 2-Methylnaphthalene	11.91	142	230067	5.58	ng	99
10) Acenaphthylene	14.01	152	1487378	4.71	ng	99
11) Acenaphthene	14.44	154	210520	4.37	ng	86
12) Fluorene	15.68	166	269711	4.57	ng	95
14) Hexachlorobenzene	16.95	284	55996	4.15	ng	98
15) Pentachlorophenol	17.48	266	23460	4.77	ng	98
16) Phenanthrene	17.95	178	387644	4.31	ng	100
17) Anthracene	18.06	178	374759	4.66	ng	99
18) Fluoranthene	20.79	202	413594	5.01	ng	99
20) Pyrene	21.31	202	432647	4.06	ng	98
22) Benzo(a)anthracene	24.21	228	374419	4.66	ng	97
23) Chrysene	24.29	228	383430	4.42	ng	99
24) Indeno(1,2,3-cd)pyrene	29.35	276	359434m	6.88	ng	
26) Benzo(b)fluoranthene	26.60	252	346026	5.69	ng	98
27) Benzo(k)fluoranthene	26.66	252	408175	5.00	ng	# 94
28) Benzo(a)pyrene	27.26	252	320525	5.61	ng	# 89
29) Dibenzo(a,h)anthracene	29.41	278	280352	6.81	ng	# 78
30) Benzo(g,h,i)perylene	29.79	276	315496	5.39	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032615\
Data File : BE089866.D
Acq On : 26 Mar 2015 20:53
Operator : TP/JJ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Quant Time: Mar 27 02:17:26 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 27 01:45:40 2015
Response via : Initial Calibration

