

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032615\
 Data File : BE089868.D
 Acq On : 26 Mar 2015 22:14
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDBE032615

Quant Time: Mar 27 02:42:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 27 02:24:32 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	117430	5.00	ng	0.00
4) Naphthalene-d8	10.23	136	491232	5.00	ng	0.00
8) Acenaphthene-d10	14.36	164	293248	5.00	ng	0.00
13) Phenanthrene-d10	17.89	188	544306	5.00	ng	0.00
19) Chrysene-d12	24.23	240	509221	5.00	ng	0.00
25) Perylene-d12	27.38	264	409124	5.00	ng	0.00

System Monitoring Compounds						
5) Nitrobenzene-d5	8.65	82	60717	2.07	ng	0.00
9) 2-Fluorobiphenyl	12.82	172	201243	3.04	ng	0.00
21) Terphenyl-d14	21.83	244	250796	3.61	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	2.48	88	16773	1.08	ng	93
3) n-Nitrosodimethylamine	2.91	42	22174	1.18	ng	# 72
6) Naphthalene	10.28	128	203046	2.01	ng	95
7) 2-Methylnaphthalene	11.90	142	175166	2.62	ng	99
10) Acenaphthylene	14.01	152	1390810	3.21	ng	99
11) Acenaphthene	14.44	154	199295	3.04	ng	86
12) Fluorene	15.68	166	263684	3.41	ng	95
14) Hexachlorobenzene	16.95	284	54107	3.23	ng	98
15) Pentachlorophenol	17.48	266	23158	0.12	ng	97
16) Phenanthrene	17.94	178	379095	3.46	ng	100
17) Anthracene	18.07	178	367274	3.39	ng	99
18) Fluoranthene	20.79	202	393981	3.28	ng	99
20) Pyrene	21.31	202	409641	3.37	ng	98
22) Benzo(a)anthracene	24.21	228	333545	3.45	ng	98
23) Chrysene	24.29	228	336213	3.46	ng	99
24) Indeno(1,2,3-cd)pyrene	29.35	276	260583m	2.83	ng	
26) Benzo(b)fluoranthene	26.60	252	277595	3.29	ng	98
27) Benzo(k)fluoranthene	26.66	252	328163	3.29	ng	# 93
28) Benzo(a)pyrene	27.26	252	251410	3.20	ng	# 89
29) Dibenzo(a,h)anthracene	29.41	278	202163	3.27	ng	# 78
30) Benzo(g,h,i)perylene	29.79	276	224941	3.12	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032615\
Data File : BE089868.D
Acq On : 26 Mar 2015 22:14
Operator : TP/JJ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDBE032615

Quant Time: Mar 27 02:42:45 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 27 02:24:32 2015
Response via : Initial Calibration

