

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032615\
 Data File : BE089857.D
 Acq On : 26 Mar 2015 13:23
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 27 03:17:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	93	0.00
2	1,4-Dioxane	1.000	1.103	-10.3	105	-0.01
3	n-Nitrosodimethylamine	1.000	0.878	12.2	85	-0.05
4 I	Naphthalene-d8	5.000	5.000	0.0	81	0.00
5 S	Nitrobenzene-d5	1.000	1.050	-5.0	100	0.00
6	Naphthalene	1.000	1.051	-5.1	88	0.00
7	2-Methylnaphthalene	1.000	1.089	-8.9	85	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	69	0.00
9 S	2-Fluorobiphenyl	1.000	1.141	-14.1	81	0.00
10	Acenaphthylene	1.000	1.120	-12.0	80	0.00
11	Acenaphthene	1.000	1.073	-7.3	78	0.00
12	Fluorene	1.000	1.043	-4.3	75	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	63	0.00
14	Hexachlorobenzene	1.000	1.109	-10.9	73	0.00
15	Pentachlorophenol	1.000	1.708	-70.8#	128	0.00
16	Phenanthrene	1.000	1.040	-4.0	70	0.00
17	Anthracene	1.000	1.063	-6.3	70	0.00
18	Fluoranthene	1.000	0.957	4.3	65	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	59	0.00
20	Pyrene	1.000	1.003	-0.3	62	0.00
21 S	Terphenyl-d14	1.000	1.094	-9.4	67	0.00
22	Benzo(a)anthracene	1.000	1.086	-8.6	69	0.00
23	Chrysene	1.000	1.063	-6.3	64	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	1.157	-15.7	93	0.00
25 I	Perylene-d12	5.000	5.000	0.0	65	0.00
26	Benzo(b)fluoranthene	1.000	1.115	-11.5	87	0.00
27	Benzo(k)fluoranthene	1.000	1.150	-15.0	75	0.00
28 C	Benzo(a)pyrene	1.000	1.123	-12.3	86	0.00
29	Dibenzo(a,h)anthracene	1.000	1.166	-16.6	87	0.01
30	Benzo(g,h,i)perylene	1.000	1.154	-15.4	83	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0