

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032615\
 Data File : BE089857.D
 Acq On : 26 Mar 2015 13:23
 Operator : TP/JJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Mar 27 03:17:38 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 25 16:20:31 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.689	0.712	-3.3	105	-0.01
3	n-Nitrosodimethylamine	0.951	0.835	12.2	85	-0.05
4 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
5 S	Nitrobenzene-d5	0.284	0.298	-4.9	100	0.00
6	Naphthalene	0.988	1.038	-5.1	88	0.00
7	2-Methylnaphthalene	0.549	0.598	-8.9	85	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	69	0.00
9 S	2-Fluorobiphenyl	1.170	1.334	-14.0	81	0.00
10	Acenaphthylene	7.154	8.009	-12.0	80	0.00
11	Acenaphthene	1.095	1.175	-7.3	78	0.00
12	Fluorene	1.328	1.386	-4.4	75	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	63	0.00
14	Hexachlorobenzene	0.166	0.184	-10.8	73	0.00
15	Pentachlorophenol	0.053	0.070	-32.1#	128	0.00
16	Phenanthrene	1.099	1.143	-4.0	70	0.00
17	Anthracene	0.982	1.045	-6.4	70	0.00
18	Fluoranthene	0.991	0.948	4.3	65	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	59	0.00
20	Pyrene	1.268	1.273	-0.4	62	0.00
21 S	Terphenyl-d14	0.738	0.807	-9.3	67	0.00
22	Benzo(a)anthracene	0.935	1.016	-8.7	69	0.00
23	Chrysene	1.020	1.084	-6.3	64	0.00
24	Indeno(1,2,3-cd)pyrene	0.578	0.800	-38.4#	93	0.00
25 I	Perylene-d12	1.000	1.000	0.0	65	0.00
26	Benzo(b)fluoranthene	0.745	0.902	-21.1	87	0.00
27	Benzo(k)fluoranthene	1.024	1.178	-15.0	75	0.00
28 C	Benzo(a)pyrene	0.702	0.868	-23.6#	86	0.00
29	Dibenzo(a,h)anthracene	0.513	0.684	-33.3#	87	0.01
30	Benzo(g,h,i)perylene	0.756	0.982	-29.9#	83	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1