

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032615\
 Data File : BE089868.D
 Acq On : 26 Mar 2015 22:14
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDBE032615

Quant Time: Mar 27 02:42:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 27 02:24:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	137	0.00
2	1,4-Dioxane	1.000	1.081	-8.1	155	0.02
3	n-Nitrosodimethylamine	1.000	1.184	-18.4	173	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	133	0.00
5 S	Nitrobenzene-d5	1.000	2.071	-107.1#	291	0.00
6	Naphthalene	1.000	2.006	-100.6#	258	0.00
7	2-Methylnaphthalene	1.000	2.618	-161.8#	340	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	140	0.00
9 S	2-Fluorobiphenyl	1.000	3.039	-203.9#	384	0.00
10	Acenaphthylene	1.000	3.210	-221.0#	436	0.00
11	Acenaphthene	1.000	3.044	-204.4#	419	0.00
12	Fluorene	1.000	3.405	-240.5#	419	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	136	0.00
14	Hexachlorobenzene	1.000	3.228	-222.8#	414	0.00
15	Pentachlorophenol	1.000	0.117	88.3#	349	0.00
16	Phenanthrene	1.000	3.462	-246.2#	417	0.00
17	Anthracene	1.000	3.387	-238.7#	440	0.00
18	Fluoranthene	1.000	3.279	-227.9#	426	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	123	0.00
20	Pyrene	1.000	3.374	-237.4#	418	0.00
21 S	Terphenyl-d14	1.000	3.612	-261.2#	412	0.00
22	Benzo(a)anthracene	1.000	3.452	-245.2#	397	0.00
23	Chrysene	1.000	3.458	-245.8#	380	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	2.831	-183.1#	348	0.00
25 I	Perylene-d12	5.000	5.000	0.0	106	0.00
26	Benzo(b)fluoranthene	1.000	3.293	-229.3#	314	0.00
27	Benzo(k)fluoranthene	1.000	3.295	-229.5#	290	0.00
28 C	Benzo(a)pyrene	1.000	3.201	-220.1#	301	0.00
29	Dibenzo(a,h)anthracene	1.000	3.271	-227.1#	346	0.00
30	Benzo(g,h,i)perylene	1.000	3.115	-211.5#	319	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1