

Data Path : Z:\HPCHEM1\BNA_E\Data\BE032615\
 Data File : BE089868.D
 Acq On : 26 Mar 2015 22:14
 Operator : TP/JJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDBE032615

Quant Time: Mar 27 02:42:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 27 02:24:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	137	0.00
2	1,4-Dioxane	0.661	0.714	-8.0	155#	0.02
3	n-Nitrosodimethylamine	0.797	0.944	-18.4	173#	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
5 S	Nitrobenzene-d5	0.298	0.618	-107.4#	291#	0.00
6	Naphthalene	1.030	2.067	-100.7#	258#	0.00
7	2-Methylnaphthalene	0.681	1.783	-161.8#	340#	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
9 S	2-Fluorobiphenyl	1.331	3.431	-157.8#	384#	0.00
10	Acenaphthylene	7.388	23.714	-221.0#	436#	0.00
11	Acenaphthene	1.116	3.398	-204.5#	419#	0.00
12	Fluorene	1.478	4.496	-204.2#	419#	0.00
13 I	Phenanthrene-d10	1.000	1.000	0.0	136	0.00
14	Hexachlorobenzene	0.154	0.497	-222.7#	414#	0.00
15	Pentachlorophenol	1.812	0.213	88.2#	349#	0.00
16	Phenanthrene	1.128	3.482	-208.7#	417#	0.00
17	Anthracene	0.996	3.374	-238.8#	440#	0.00
18	Fluoranthene	1.104	3.619	-227.8#	426#	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
20	Pyrene	1.192	4.022	-237.4#	418#	0.00
21 S	Terphenyl-d14	0.811	2.463	-203.7#	412#	0.00
22	Benzo(a)anthracene	1.093	3.275	-199.6#	397#	0.00
23	Chrysene	1.110	3.301	-197.4#	380#	0.00
24	Indeno(1,2,3-cd)pyrene	0.970	2.559	-163.8#	348#	0.00
25 I	Perylene-d12	1.000	1.000	0.0	106	0.00
26	Benzo(b)fluoranthene	2.154	3.393	-57.5#	314#	0.00
27	Benzo(k)fluoranthene	2.824	4.011	-42.0#	290#	0.00
28 C	Benzo(a)pyrene	1.849	3.073	-66.2#	301#	0.00
29	Dibenzo(a,h)anthracene	0.755	2.471	-227.3#	346#	0.00
30	Benzo(g,h,i)perylene	0.990	2.749	-177.7#	319#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1