


Data Path : Z:\HPCHEM1\BNA_E\Data\BE032615\
 Data File : BE089859.D
 Acq On : 26 Mar 2015 16:07
 Operator : TP/JJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE032415.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Mar 25 16:20:31 2015



AutoFind: Scans 3232, 3233, 3234; Background Corrected with Scan 3216

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.8	191920	PASS
68	69	0.00	2	1.4	2980	PASS
69	198	0.00	100	38.3	211215	PASS
70	69	0.00	2	0.3	664	PASS
127	198	10	80	54.7	301526	PASS
197	198	0.00	2	0.1	625	PASS
198	198	100	100	100.0	551210	PASS
199	198	5	9	6.4	35142	PASS
275	198	10	60	20.1	110538	PASS
365	198	1	100	2.5	13729	PASS
441	442	0.01	24	19.5	66737	PASS
442	198	50	100	62.2	342848	PASS
443	442	15	24	19.4	66350	PASS