

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032618\
 Data File : BE095857.D
 Acq On : 27 Mar 2018 2:20
 Operator : SJ/JU
 Sample : J1994-06
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C-MW-06-032018

Manual Integrations
 APPROVED

Sohil
 3/27/2018 4:58:18 PM

Quant Time: Mar 27 14:48:34 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1222	0.40	ng	0.00
7) Naphthalene-d8	11.26	136	4663	0.40	ng	0.00
13) Acenaphthene-d10	14.99	164	2775	0.40	ng	-0.01
19) Phenanthrene-d10	17.70	188	6429	0.40	ng	-0.01
27) Chrysene-d12	21.79	240	8594	0.40	ng	-0.01
34) Perylene-d12	24.42	264	7023	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.91	112	644	0.16	ng	0.00
5) Phenol-d6	7.56	99	596	0.11	ng	0.00
8) Nitrobenzene-d5	9.59	82	2377	0.43	ng	0.00
11) 2-Methylnaphthalene-d10	12.79	152	2806	0.36	ng	-0.01
14) 2,4,6-Tribromophenol	16.46	330	588	0.37	ng	-0.01
15) 2-Fluorobiphenyl	13.63	172	6121	0.52	ng	-0.01
25) Fluoranthene-d10	19.68	212	33606	0.35	ng	-0.01
29) Terphenyl-d14	20.24	244	7801	0.43	ng	0.00
Target Compounds						Qvalue
23) Phenanthrene	17.74	178	637	0.033	ng	99
26) Fluoranthene	19.71	202	817	0.031	ng	94
28) Pyrene	20.06	202	1132m	0.032	ng	
32) Bis(2-ethylhexyl)phthalate	21.64	149	24973	0.274	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032618\
Data File : BE095857.D
Acq On : 27 Mar 2018 2:20
Operator : SJ/JU
Sample : J1994-06
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C-MW-06-032018

Manual Integrations
APPROVED

Sohil
3/27/2018 4:58:18 PM

Quant Time: Mar 27 14:48:34 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

