

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032618\
 Data File : BE095860.D
 Acq On : 27 Mar 2018 4:07
 Operator : SJ/JU
 Sample : J1994-12
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C-MW-09-031918

Quant Time: Mar 27 07:02:22 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1179	0.40	ng	-0.01
7) Naphthalene-d8	11.26	136	4417	0.40	ng	0.00
13) Acenaphthene-d10	14.99	164	2623	0.40	ng	-0.01
19) Phenanthrene-d10	17.70	188	6176	0.40	ng	-0.01
27) Chrysene-d12	21.79	240	8956	0.40	ng	-0.01
34) Perylene-d12	24.42	264	6993	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.91	112	694	0.18	ng	-0.01
5) Phenol-d6	7.56	99	619	0.12	ng	0.00
8) Nitrobenzene-d5	9.59	82	2410	0.46	ng	0.00
11) 2-Methylnaphthalene-d10	12.79	152	2755	0.37	ng	0.00
14) 2,4,6-Tribromophenol	16.46	330	594	0.39	ng	-0.01
15) 2-Fluorobiphenyl	13.63	172	6009	0.54	ng	-0.01
25) Fluoranthene-d10	19.68	212	33997	0.37	ng	0.00
29) Terphenyl-d14	20.24	244	8812	0.46	ng	0.00
Target Compounds						
23) Phenanthrene	17.74	178	722	0.039	ng	Qvalue 96
26) Fluoranthene	19.71	202	1620	0.063	ng	99
28) Pyrene	20.06	202	1836	0.050	ng	97
32) Bis(2-ethylhexyl)phthalate	21.64	149	6955	0.073	ng	# 98
33) Indeno(1,2,3-cd)pyrene	27.22	276	776	0.025	ng	# 67
35) Benzo(b)fluoranthene	23.60	252	1507	0.065	ng	# 60
39) Benzo(g,h,i)perylene	28.09	276	788	0.038	ng	# 54

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032618\
Data File : BE095860.D
Acq On : 27 Mar 2018 4:07
Operator : SJ/JU
Sample : J1994-12
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C-MW-09-031918

Quant Time: Mar 27 07:02:22 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

