

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032618\
 Data File : BE095863.D
 Acq On : 27 Mar 2018 5:55
 Operator : SJ/JU
 Sample : J1994-15
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PR-MW-01-031918

Manual Integrations
 APPROVED

Sohil
 3/27/2018 4:58:21 PM

Quant Time: Mar 27 15:33:21 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1426	0.40	ng	0.00
7) Naphthalene-d8	11.26	136	5596	0.40	ng	0.00
13) Acenaphthene-d10	15.01	164	6022m	0.40	ng	0.01
19) Phenanthrene-d10	17.71	188	6395m	0.40	ng	0.00
27) Chrysene-d12	21.79	240	7572	0.40	ng	-0.01
34) Perylene-d12	24.42	264	7501	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.91	112	1255	0.27	ng	0.00
5) Phenol-d6	7.57	99	762	0.12	ng	0.00
8) Nitrobenzene-d5	9.59	82	5636m	0.85	ng	0.00
11) 2-Methylnaphthalene-d10	12.84	152	4369	0.47	ng	0.04
14) 2,4,6-Tribromophenol	16.47	330	673	0.19	ng	0.00
15) 2-Fluorobiphenyl	13.64	172	6042	0.24	ng	0.00
25) Fluoranthene-d10	19.69	212	28885	0.30	ng	0.00
29) Terphenyl-d14	20.24	244	9039	0.56	ng	0.00
Target Compounds						
9) Naphthalene	11.31	128	3337813	52.222	ng	99
12) 2-Methylnaphthalene	12.87	142	407836	37.362	ng	93
17) Acenaphthene	15.06	154	4231	0.201	ng	# 55
18) Fluorene	16.03	166	11289	0.409	ng	# 79
22) Pentachlorophenol	17.37	266	330	0.359	ng	85
32) Bis(2-ethylhexyl)phthalate	21.64	149	4704	0.059	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032618\
Data File : BE095863.D
Acq On : 27 Mar 2018 5:55
Operator : SJ/JU
Sample : J1994-15
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PR-MW-01-031918

Manual Integrations
APPROVED

Sohil
3/27/2018 4:58:21 PM

Quant Time: Mar 27 15:33:21 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

