

Data Path : Z:\HPCHEM1\BNA E\Data\BE032618\
 Data File : BE095846.D
 Acq On : 26 Mar 2018 18:37
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 05:42:28 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.655	0.628	4.1	103	0.00
3	n-Nitrosodimethylamine	0.961	0.935	2.7	96	0.00
4 S	2-Fluorophenol	1.319	1.270	3.7	94	0.00
5 S	Phenol-d6	1.813	1.804	0.5	97	0.00
6	bis(2-Chloroethyl)ether	1.564	1.596	-2.0	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.473	0.471	0.4	99	0.00
9	Naphthalene	4.569	4.649	-1.8	98	-0.01
10	Hexachlorobutadiene	0.314	0.306	2.5	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.668	0.682	-2.1	99	-0.01
12	2-Methylnaphthalene	0.780	0.797	-2.2	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.231	0.231	0.0	102	-0.01
15 S	2-Fluorobiphenyl	1.686	1.701	-0.9	98	-0.01
16	Acenaphthylene	2.288	2.276	0.5	98	0.00
17	Acenaphthene	1.395	1.398	-0.2	100	0.00
18	Fluorene	1.835	1.836	-0.1	98	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	-0.01
20	4-Bromophenyl-phenylether	0.240	0.231	3.7	97	0.00
21	Hexachlorobenzene	0.240	0.237	1.3	99	0.00
22	Pentachlorophenol	0.069	0.070	-1.4	125	-0.01
23	Phenanthrene	1.197	1.214	-1.4	101	0.00
24	Anthracene	1.197	1.173	2.0	99	0.00
25 SURR	Fluoranthene-d10	5.940	5.938	0.0	100	0.00
26	Fluoranthene	1.663	1.675	-0.7	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	-0.01
28	Pvrene	1.629	1.624	0.3	93	-0.01
29 S	Terphenyl-d14	0.850	0.831	2.2	99	0.00
30	Benzo(a)anthracene	1.387	1.405	-1.3	101	0.00
31	Chrysene	1.266	1.293	-2.1	103	-0.01
32	Bis(2-ethylhexyl)phthalate	4.242	3.895	8.2	100	0.00
33	Indeno(1,2,3-cd)pyrene	1.364	1.360	0.3	101	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.02
35	Benzo(b)fluoranthene	1.323	1.383	-4.5	103	-0.01
36	Benzo(k)fluoranthene	1.324	1.376	-3.9	103	-0.01
37 C	Benzo(a)pyrene	1.240	1.276	-2.9	101	-0.01
38	Dibenzo(a,h)anthracene	1.178	1.217	-3.3	101	-0.03
39	Benzo(a,h,i)perylene	1.184	1.235	-4.3	103	-0.02

Data Path : Z:\HPCHEM1\BNA E\Data\BE032618\
Data File : BE095846.D
Acq On : 26 Mar 2018 18:37
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 05:42:28 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			