

Data Path : Z:\HPCHEM1\BNA E\Data\BE032618\
 Data File : BE095848.D
 Acq On : 26 Mar 2018 20:25
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 05:42:48 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.655	0.640	2.3	107	0.00
3	n-Nitrosodimethylamine	0.961	0.877	8.7	92	0.00
4 S	2-Fluorophenol	1.319	1.273	3.5	96	0.00
5 S	Phenol-d6	1.813	1.784	1.6	98	0.00
6	bis(2-Chloroethyl)ether	1.564	1.533	2.0	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.473	0.470	0.6	100	0.00
9	Naphthalene	4.569	4.678	-2.4	100	-0.01
10	Hexachlorobutadiene	0.314	0.302	3.8	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.668	0.669	-0.1	98	-0.01
12	2-Methylnaphthalene	0.780	0.804	-3.1	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	-0.01
14 S	2,4,6-Tribromophenol	0.231	0.216	6.5	98	-0.01
15 S	2-Fluorobiphenyl	1.686	1.704	-1.1	100	-0.01
16	Acenaphthylene	2.288	2.314	-1.1	102	0.00
17	Acenaphthene	1.395	1.380	1.1	101	0.00
18	Fluorene	1.835	1.883	-2.6	102	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	-0.01
20	4-Bromophenyl-phenylether	0.240	0.242	-0.8	101	0.00
21	Hexachlorobenzene	0.240	0.241	-0.4	101	0.00
22	Pentachlorophenol	0.069	0.065	5.8	114	-0.01
23	Phenanthrene	1.197	1.259	-5.2	104	0.00
24	Anthracene	1.197	1.213	-1.3	102	0.00
25 SURR	Fluoranthene-d10	5.940	5.995	-0.9	101	0.00
26	Fluoranthene	1.663	1.708	-2.7	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	-0.01
28	Pvrene	1.629	1.547	5.0	88	-0.01
29 S	Terphenyl-d14	0.850	0.848	0.2	100	0.00
30	Benzo(a)anthracene	1.387	1.434	-3.4	103	-0.01
31	Chrysene	1.266	1.319	-4.2	104	-0.01
32	Bis(2-ethylhexyl)phthalate	4.242	3.854	9.1	98	-0.01
33	Indeno(1,2,3-cd)pyrene	1.364	1.383	-1.4	102	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	101	-0.02
35	Benzo(b)fluoranthene	1.323	1.370	-3.6	103	-0.01
36	Benzo(k)fluoranthene	1.324	1.374	-3.8	104	-0.01
37 C	Benzo(a)pyrene	1.240	1.278	-3.1	102	-0.02
38	Dibenzo(a,h)anthracene	1.178	1.214	-3.1	101	-0.02
39	Benzo(a,h,i)perylene	1.184	1.221	-3.1	102	-0.03

Data Path : Z:\HPCHEM1\BNA E\Data\BE032618\
Data File : BE095848.D
Acq On : 26 Mar 2018 20:25
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 05:42:48 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			