

Data Path : Z:\HPCHEM1\BNA E\Data\BE032618\
 Data File : BE095864.D
 Acq On : 27 Mar 2018 6:31
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 07:15:50 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.403	-0.8	113	0.00
3	n-Nitrosodimethylamine	0.400	0.379	5.3	97	0.00
4 S	2-Fluorophenol	0.400	0.391	2.3	100	0.00
5 S	Phenol-d6	0.400	0.411	-2.7	104	0.00
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.00
8 S	Nitrobenzene-d5	0.400	0.393	1.8	105	0.00
9	Naphthalene	0.400	0.408	-2.0	107	-0.01
10	Hexachlorobutadiene	0.400	0.417	-4.2	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.408	-2.0	107	-0.01
12	2-Methylnaphthalene	0.400	0.420	-5.0	110	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.440	-10.0	125	-0.01
15 S	2-Fluorobiphenyl	0.400	0.401	-0.3	108	-0.01
16	Acenaphthylene	0.400	0.409	-2.2	113	0.00
17	Acenaphthene	0.400	0.398	0.5	111	0.00
18	Fluorene	0.400	0.405	-1.3	110	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	-0.01
20	4-Bromophenyl-phenylether	0.400	0.420	-5.0	113	0.00
21	Hexachlorobenzene	0.400	0.411	-2.7	110	0.00
22	Pentachlorophenol	0.400	0.608	-52.0#	204	-0.01
23	Phenanthrene	0.400	0.415	-3.7	110	0.00
24	Anthracene	0.400	0.415	-3.7	112	0.00
25 SURR	Fluoranthene-d10	0.400	0.399	0.3	107	0.00
26	Fluoranthene	0.400	0.403	-0.8	109	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	103	-0.01
28	Pvrene	0.400	0.473	-18.2	110	-0.01
29 S	Terphenyl-d14	0.400	0.411	-2.7	105	0.00
30	Benzo(a)anthracene	0.400	0.411	-2.7	103	0.00
31	Chrysene	0.400	0.417	-4.2	106	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.374	6.5	103	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.412	-3.0	104	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	103	-0.02
35	Benzo(b)fluoranthene	0.400	0.418	-4.5	106	-0.01
36	Benzo(k)fluoranthene	0.400	0.406	-1.5	104	-0.01
37 C	Benzo(a)pyrene	0.400	0.411	-2.7	104	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.417	-4.2	104	-0.03
39	Benzo(a,h,i)perylene	0.400	0.414	-3.5	104	-0.03

Data Path : Z:\HPCHEM1\BNA E\Data\BE032618\
Data File : BE095864.D
Acq On : 27 Mar 2018 6:31
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 07:15:50 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				