

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032718\
 Data File : BE095866.D
 Acq On : 27 Mar 2018 9:39
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 17:38:48 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	-0.01
2	1,4-Dioxane	0.400	0.417	-4.2	117	0.00
3	n-Nitrosodimethylamine	0.400	0.381	4.8	98	0.00
4 S	2-Fluorophenol	0.400	0.403	-0.8	103	-0.01
5 S	Phenol-d6	0.400	0.418	-4.5	106	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.415	-3.7	106	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	109	0.00
8 S	Nitrobenzene-d5	0.400	0.394	1.5	106	0.00
9	Naphthalene	0.400	0.413	-3.2	109	-0.01
10	Hexachlorobutadiene	0.400	0.410	-2.5	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	108	0.00
12	2-Methylnaphthalene	0.400	0.419	-4.7	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	112	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.431	-7.7	124	-0.01
15 S	2-Fluorobiphenyl	0.400	0.394	1.5	108	-0.01
16	Acenaphthylene	0.400	0.398	0.5	111	0.00
17	Acenaphthene	0.400	0.404	-1.0	114	0.00
18	Fluorene	0.400	0.405	-1.3	112	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	-0.01
20	4-Bromophenyl-phenylether	0.400	0.418	-4.5	114	0.00
21	Hexachlorobenzene	0.400	0.410	-2.5	112	0.00
22	Pentachlorophenol	0.400	0.586	-46.5#	199	-0.01
23	Phenanthrene	0.400	0.418	-4.5	113	0.00
24	Anthracene	0.400	0.404	-1.0	111	-0.01
25 SURR	Fluoranthene-d10	0.400	0.401	-0.3	109	0.00
26	Fluoranthene	0.400	0.405	-1.3	111	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	107	-0.01
28	Pvrene	0.400	0.421	-5.2	102	-0.01
29 S	Terphenyl-d14	0.400	0.407	-1.7	108	0.00
30	Benzo(a)anthracene	0.400	0.404	-1.0	106	-0.01
31	Chrysene	0.400	0.413	-3.2	109	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.359	10.3	102	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.407	-1.7	107	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	105	-0.02
35	Benzo(b)fluoranthene	0.400	0.417	-4.2	108	-0.01
36	Benzo(k)fluoranthene	0.400	0.424	-6.0	110	-0.01
37 C	Benzo(a)pyrene	0.400	0.418	-4.5	108	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.420	-5.0	107	-0.03
39	Benzo(a,h,i)perylene	0.400	0.416	-4.0	107	-0.03

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032718\
Data File : BE095866.D
Acq On : 27 Mar 2018 9:39
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 17:38:48 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				