

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032718\
 Data File : BE095866.D
 Acq On : 27 Mar 2018 9:39
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 17:38:48 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	-0.01
2	1,4-Dioxane	0.655	0.683	-4.3	117	0.00
3	n-Nitrosodimethylamine	0.961	0.916	4.7	98	0.00
4 S	2-Fluorophenol	1.319	1.328	-0.7	103	-0.01
5 S	Phenol-d6	1.813	1.894	-4.5	106	-0.01
6	bis(2-Chloroethyl)ether	1.564	1.624	-3.8	106	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.473	0.466	1.5	106	0.00
9	Naphthalene	4.569	4.715	-3.2	109	-0.01
10	Hexachlorobutadiene	0.314	0.321	-2.2	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.668	0.684	-2.4	108	0.00
12	2-Methylnaphthalene	0.780	0.818	-4.9	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	112	-0.01
14 S	2,4,6-Tribromophenol	0.231	0.249	-7.8	124	-0.01
15 S	2-Fluorobiphenyl	1.686	1.660	1.5	108	-0.01
16	Acenaphthylene	2.288	2.276	0.5	111	0.00
17	Acenaphthene	1.395	1.411	-1.1	114	0.00
18	Fluorene	1.835	1.857	-1.2	112	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	-0.01
20	4-Bromophenyl-phenylether	0.240	0.251	-4.6	114	0.00
21	Hexachlorobenzene	0.240	0.246	-2.5	112	0.00
22	Pentachlorophenol	0.069	0.103	-49.3#	199#	-0.01
23	Phenanthrene	1.197	1.252	-4.6	113	0.00
24	Anthracene	1.197	1.209	-1.0	111	-0.01
25 SURR	Fluoranthene-d10	5.940	5.960	-0.3	109	0.00
26	Fluoranthene	1.663	1.684	-1.3	111	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	-0.01
28	Pvrene	1.629	1.715	-5.3	102	-0.01
29 S	Terphenyl-d14	0.850	0.865	-1.8	108	0.00
30	Benzo(a)anthracene	1.387	1.403	-1.2	106	-0.01
31	Chrysene	1.266	1.306	-3.2	109	-0.01
32	Bis(2-ethylhexyl)phthalate	4.242	3.812	10.1	102	0.00
33	Indeno(1,2,3-cd)pyrene	1.364	1.387	-1.7	107	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	105	-0.02
35	Benzo(b)fluoranthene	1.323	1.379	-4.2	108	-0.01
36	Benzo(k)fluoranthene	1.324	1.404	-6.0	110	-0.01
37 C	Benzo(a)pyrene	1.240	1.295	-4.4	108	-0.02
38	Dibenzo(a,h)anthracene	1.178	1.239	-5.2	107	-0.03
39	Benzo(a,h,i)perylene	1.184	1.232	-4.1	107	-0.03

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032718\
Data File : BE095866.D
Acq On : 27 Mar 2018 9:39
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 17:38:48 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			