

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032816\
 Data File : BE091532.D
 Acq On : 28 Mar 2016 9:56
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 28 19:15:20 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:00:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.669	0.635	5.1	94	0.00
3	n-Nitrosodimethylamine	0.581	0.542	6.7	98	0.00
4 S	2-Fluorophenol	0.975	0.894	8.3	102	0.00
5 S	Phenol-d6	1.330	1.193	10.3	104	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
7 S	Nitrobenzene-d5	0.217	0.183	15.7	101	-0.01
8	Naphthalene	1.007	0.997	1.0	102	0.00
9	2-Methylnaphthalene	0.629	0.603	4.1	103	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	101	-0.01
11 S	2,4,6-Tribromophenol	0.085	0.071	16.5	99	-0.01
12 S	2-Fluorobiphenyl	1.408	1.394	1.0	102	-0.01
13	Acenaphthylene	1.752	1.547	11.7	101	0.00
14	Acenaphthene	1.237	1.245	-0.6	102	-0.01
15	Fluorene	1.511	1.444	4.4	102	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
17	Hexachlorobenzene	0.176	0.172	2.3	105	0.00
18	Pentachlorophenol	0.047	0.041	12.8	108	-0.02
19	Phenanthrene	1.145	1.146	-0.1	103	0.00
20	Anthracene	0.948	0.861	9.2	103	0.00
21	Fluoranthene	1.071	1.009	5.8	106	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
23	Pyrene	1.181	1.044	11.6	103	0.00
24 S	Terphenyl-d14	0.822	0.770	6.3	107	0.00
25	Benzo(a)anthracene	1.154	1.070	7.3	108	0.00
26	Chrysene	1.516	1.522	-0.4	106	0.00
27	Indeno(1,2,3-cd)pyrene	1.179	1.081	8.3	110	-0.01
28 I	Perylene-d12	1.000	1.000	0.0	107	-0.01
29	Benzo(b)fluoranthene	1.111	1.090	1.9	119	0.00
30	Benzo(k)fluoranthene	1.107	0.954	13.8	96	0.00
31 C	Benzo(a)pyrene	0.926	0.807	12.9	110	0.00
32	Dibenzo(a,h)anthracene	0.984	0.888	9.8	108	-0.01
33	Benzo(g,h,i)perylene	1.029	0.946	8.1	108	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0