

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032816\
 Data File : BE091532.D
 Acq On : 28 Mar 2016 9:56
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 28 19:15:20 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:00:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	104	0.00
2	1,4-Dioxane	0.400	0.376	6.0	94	0.00
3	n-Nitrosodimethylamine	0.400	0.382	4.5	98	0.00
4 S	2-Fluorophenol	0.400	0.367	8.3	102	0.00
5 S	Phenol-d6	0.400	0.359	10.3	104	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	102	0.00
7 S	Nitrobenzene-d5	0.400	0.338	15.5	101	-0.01
8	Naphthalene	0.400	0.396	1.0	102	0.00
9	2-Methylnaphthalene	0.400	0.383	4.3	103	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	101	-0.01
11 S	2,4,6-Tribromophenol	0.400	0.334	16.5	99	-0.01
12 S	2-Fluorobiphenyl	0.400	0.396	1.0	102	-0.01
13	Acenaphthylene	0.400	0.353	11.8	101	0.00
14	Acenaphthene	0.400	0.403	-0.8	102	-0.01
15	Fluorene	0.400	0.382	4.5	102	-0.01
16 I	Phenanthrene-d10	5.000	5.000	0.0	103	0.00
17	Hexachlorobenzene	0.400	0.393	1.8	105	0.00
18	Pentachlorophenol	0.400	0.438	-9.5	108	-0.02
19	Phenanthrene	0.400	0.400	0.0	103	0.00
20	Anthracene	0.400	0.363	9.3	103	0.00
21	Fluoranthene	0.400	0.377	5.8	106	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	105	0.00
23	Pyrene	0.400	0.354	11.5	103	0.00
24 S	Terphenyl-d14	0.400	0.375	6.3	107	0.00
25	Benzo(a)anthracene	0.400	0.371	7.3	108	0.00
26	Chrysene	0.400	0.402	-0.5	106	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.367	8.3	110	-0.01
28 I	Perylene-d12	5.000	5.000	0.0	107	-0.01
29	Benzo(b)fluoranthene	0.400	0.393	1.8	119	0.00
30	Benzo(k)fluoranthene	0.400	0.345	13.8	96	0.00
31 C	Benzo(a)pyrene	0.400	0.483	-20.7#	110	0.00
32	Dibenzo(a,h)anthracene	0.400	0.361	9.8	108	-0.01
33	Benzo(g,h,i)perylene	0.400	0.368	8.0	108	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1