

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032816\
 Data File : BE091543.D
 Acq On : 28 Mar 2016 18:33
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 01:22:14 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 25 19:00:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	86	0.00
2	1,4-Dioxane	0.400	0.410	-2.5	83	-0.02
3	n-Nitrosodimethylamine	0.400	0.405	-1.3	87	0.00
4 S	2-Fluorophenol	0.400	0.380	5.0	87	-0.02
5 S	Phenol-d6	0.400	0.363	9.3	86	-0.02
6 I	Naphthalene-d8	5.000	5.000	0.0	87	0.00
7 S	Nitrobenzene-d5	0.400	0.373	6.8	94	-0.01
8	Naphthalene	0.400	0.395	1.3	87	-0.02
9	2-Methylnaphthalene	0.400	0.382	4.5	88	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	85	-0.01
11 S	2,4,6-Tribromophenol	0.400	0.390	2.5	97	-0.01
12 S	2-Fluorobiphenyl	0.400	0.396	1.0	86	-0.01
13	Acenaphthylene	0.400	0.368	8.0	88	0.00
14	Acenaphthene	0.400	0.402	-0.5	86	-0.01
15	Fluorene	0.400	0.381	4.8	86	-0.01
16 I	Phenanthrene-d10	5.000	5.000	0.0	87	0.00
17	Hexachlorobenzene	0.400	0.383	4.3	88	0.00
18	Pentachlorophenol	0.400	0.441	-10.2	93	-0.02
19	Phenanthrene	0.400	0.394	1.5	86	0.00
20	Anthracene	0.400	0.366	8.5	89	0.00
21	Fluoranthene	0.400	0.370	7.5	88	-0.02
22 I	Chrysene-d12	5.000	5.000	0.0	90	0.00
23	Pyrene	0.400	0.357	10.8	89	0.00
24 S	Terphenyl-d14	0.400	0.361	9.8	88	0.00
25	Benzo(a)anthracene	0.400	0.356	11.0	88	0.00
26	Chrysene	0.400	0.374	6.5	84	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.355	11.3	90	-0.01
28 I	Perylene-d12	5.000	5.000	0.0	88	-0.01
29	Benzo(b)fluoranthene	0.400	0.397	0.8	99	0.00
30	Benzo(k)fluoranthene	0.400	0.345	13.8	80	0.00
31 C	Benzo(a)pyrene	0.400	0.487	-21.7#	92	-0.01
32	Dibenzo(a,h)anthracene	0.400	0.363	9.3	90	-0.01
33	Benzo(g,h,i)perylene	0.400	0.365	8.8	88	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1