


Data Path : Z:\HPCHEM1\BNA\_E\DATA\BE032816\  
 Data File : BE091542.D  
 Acq On : 28 Mar 2016 16:36  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE032516.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Mar 25 19:00:32 2016



AutoFind: Scans 2555, 2556, 2557; Background Corrected with Scan 2547

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 57.5      | 56145   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.7       | 831     | PASS             |
| 69          | 198          | 0.00         | 100          | 49.5      | 48287   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 170     | PASS             |
| 127         | 198          | 10           | 80           | 54.4      | 53125   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 97569   | PASS             |
| 199         | 198          | 5            | 9            | 6.5       | 6384    | PASS             |
| 275         | 198          | 10           | 60           | 21.9      | 21321   | PASS             |
| 365         | 198          | 1            | 100          | 2.9       | 2854    | PASS             |
| 441         | 442          | 0.01         | 24           | 18.1      | 10011   | PASS             |
| 442         | 198          | 50           | 100          | 56.6      | 55181   | PASS             |
| 443         | 442          | 15           | 24           | 19.2      | 10585   | PASS             |