

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032817\
 Data File : BE092832.D
 Acq On : 28 Mar 2017 11:59
 Operator : SJ/MA
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :


Quant Time: Mar 28 13:08:43 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 28 13:07:38 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	7414	5.00	ng	0.00
7) Naphthalene-d8	10.54	136	33634	5.00	ng	0.00
12) Acenaphthene-d10	14.40	164	16621	5.00	ng	0.00
18) Phenanthrene-d10	17.16	188	36737	5.00	ng	0.00
24) Chrysene-d12	21.30	240	42996	5.00	ng	0.00
31) Perylene-d12	23.71	264	37502	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	2656	1.69	ng	0.00
5) Phenol-d6	6.94	99	3882	2.08	ng	0.00
8) Nitrobenzene-d5	8.93	82	2890	1.41	ng	0.00
13) 2,4,6-Tribromophenol	15.88	330	392	0.90	ng	0.00
14) 2-Fluorobiphenyl	13.02	172	8861	2.26	ng	-0.01
26) Terphenyl-d14	19.76	244	11665	2.25	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.32	88	1503	1.36	ng	95
3) n-Nitrosodimethylamine	3.61	42	1609	1.29	ng	85
6) bis(2-Chloroethyl)ether	7.20	93	3905	1.92	ng	# 46
9) Naphthalene	10.58	128	11505	1.60	ng	# 96
10) Hexachlorobutadiene	10.89	225	1548	1.49	ng	99
11) 2-Methylnaphthalene	12.21	142	7382	1.59	ng	98
15) Acenaphthylene	14.11	152	40388	1.58	ng	100
16) Acenaphthene	14.45	154	7260	2.03	ng	92
17) Fluorene	15.44	166	8739	1.80	ng	99
19) Hexachlorobenzene	16.47	284	2275	1.51	ng	# 68
20) Pentachlorophenol	16.81	266	531	1.05	ng	91
21) Phenanthrene	17.18	178	14677	1.67	ng	# 94
22) Anthracene	17.27	178	12987	1.46	ng	98
23) Fluoranthene	19.18	202	63863	1.49	ng	99
25) Pyrene	19.54	202	69780	1.79	ng	99
27) Benzo(a)anthracene	21.26	228	16017	0.37	ng	96
28) Chrysene	21.31	228	17206	0.44	ng	# 86
29) Bis(2-ethylhexyl)phthalate	21.23	149	4145	0.88	ng	# 74
30) Indeno(1,2,3-cd)pyrene	26.25	276	65001	1.52	ng	# 93
32) Benzo(b)fluoranthene	22.97	252	16029	1.60	ng	99
33) Benzo(k)fluoranthene	23.02	252	14355	1.43	ng	95
34) Benzo(a)pyrene	23.60	252	13095	1.44	ng	97
35) Dibenzo(a,h)anthracene	26.29	278	14172	1.63	ng	97
36) Benzo(g,h,i)perylene	27.02	276	57753	1.56	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.80 min Scan# 556

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :


Tot Ion:152 Resp: 7414

Ion Ratio Lower Upper

152 100

150 126.7 106.0 159.0




115 33.7 31.2 46.8



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 1.36 ng

RT: 3.32 min Scan# 63

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59


Tgt Ion: 88 Resp: 1503

Ion Ratio Lower Upper

88 100

58 76.8 63.6 95.4



43 29.3 28.0 42.0




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 1.29 ng

RT: 3.61 min Scan# 88

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :


Tot Ion: 42 Resp: 1609

Ion Ratio Lower Upper




42 100

74 123.4 86.0 129.0

44 5.7 5.1 7.7



Abundance Ion 42.00 (41.70 to 42.70): BE0
Ion 74.00 (73.70 to 74.70): BE0
Ion 44.00 (43.70 to 44.70): BE0



#4

2-Fluorophenol

Concen: 1.69 ng

RT: 5.36 min Scan# 267

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59


Tgt Ion: 112 Resp: 2656

Ion Ratio Lower Upper



112 100


64 69.2 53.5 80.3

63 35.8 27.9 41.9



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BE0
Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 2.08 ng

RT: 6.94 min Scan# 519

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :


Tot Ion: 99 Resp: 3882

Ion Ratio Lower Upper




99 100

42 0.0 16.0 24.0#

71 0.0 30.6 45.8#



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 1.92 ng

RT: 7.20 min Scan# 530

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59


Tgt Ion: 93 Resp: 3905

Ion Ratio Lower Upper



93 100


63 82.0 16.0 24.0#

95 29.9 24.0 36.0



Abundance Ion 93.00 (92.70 to 93.70): BE0
Ion 63.00 (62.70 to 63.70): BE0
Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.54 min Scan# 679

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :

Tot Ion:136 Resp: 33634


Ion Ratio Lower Upper

136 100

137 11.3 8.2 12.4

54 5.8 9.6 14.4#

68 5.5 6.7 10.1#




Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0




Abundance

Time-->

10.54

10.40 10.50 10.60 10.70



#8

Nitrobenzene-d5

Concen: 1.41 ng

RT: 8.93 min Scan# 605

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59


Tgt Ion: 82 Resp: 2890

Ion Ratio Lower Upper

82 100

128 43.5 39.0 58.4


54 33.6 44.8 67.2#



Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0




Abundance

Time-->

8.93

8.70 8.80 8.90 9.00 9.10



#9

Naphthalene

Concen: 1.60 ng

RT: 10.58 min Scan# 681

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :


Tot Ion:128 Resp: 11505

Ion Ratio Lower Upper




128 100

129 10.7 11.2 16.8#

127 14.6 11.6 17.4



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E



#10

Hexachlorobutadiene

Concen: 1.49 ng

RT: 10.89 min Scan# 697

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59


Tgt Ion:225 Resp: 1548

Ion Ratio Lower Upper



225 100


223 62.7 50.6 76.0

227 63.3 51.4 77.0



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E





#11

2-Methylnaphthalene

Concen: 1.59 ng

RT: 12.21 min Scan# 770

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :





Tot Ion:142 Resp: 7382

Ion Ratio Lower Upper

142 100

141 91.3 73.7 110.5

115 38.8 34.0 51.0



#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.40 min Scan# 933

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59




Tgt Ion:164 Resp: 16621


Ion Ratio Lower Upper

164 100

162 98.2 71.4 107.0

160 42.0 27.4 41.2#









#13
 2,4,6-Tribromophenol
 Concen: 0.90 ng
 RT: 15.88 min Scan# 1042
 Delta R.T. -0.00 min
 Lab File: BE092832.D
 Acq: 28 Mar 2017 11:59




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:330 Resp: 392
 Ion Ratio Lower Upper
 330 100
 332 99.0 75.4 113.0
 141 56.4 31.0 46.6#



#14
 2-Fluorobiphenyl
 Concen: 2.26 ng
 RT: 13.02 min Scan# 840
 Delta R.T. -0.01 min
 Lab File: BE092832.D
 Acq: 28 Mar 2017 11:59

Tgt Ion:172 Resp: 8861
 Ion Ratio Lower Upper
 172 100
 171 38.6 30.1 45.1
 170 27.7 21.8 32.6









#15
Acenaphthylene
Concen: 1.58 ng
RT: 14.11 min Scan# 916
Delta R.T. -0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59




Instrument :
BNA_E
ClientSampleId :


Tot Ion:152 Resp: 40388
Ion Ratio Lower Upper
152 100
151 4.9 3.9 5.9
153 13.0 10.4 15.6



#16
Acenaphthene
Concen: 2.03 ng
RT: 14.45 min Scan# 936
Delta R.T. -0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Tgt Ion:154 Resp: 7260
Ion Ratio Lower Upper
154 100
153 460.9 350.8 526.2
152 224.9 171.1 256.7









#17
Fluorene
Concen: 1.80 ng
RT: 15.44 min Scan# 1004
Delta R.T. -0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59




Instrument :
BNA_E
ClientSampleId :


Tot Ion:166 Resp: 8739
Ion Ratio Lower Upper
166 100
165 99.1 78.2 117.4
167 14.0 11.0 16.4



#18
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.16 min Scan# 1109
Delta R.T. 0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Tgt Ion:188 Resp: 36737
Ion Ratio Lower Upper
188 100
94 3.0 3.2 4.8#
80 2.3 2.6 3.8#









#19
Hexachlorobenzene
Concen: 1.51 ng
RT: 16.47 min Scan# 1079
Delta R.T. -0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Instrument :
BNA_E
ClientSampleId :

Tot Ion:284 Resp: 2275
Ion Ratio Lower Upper
284 100
142 0.0 29.1 43.7#
249 33.8 27.9 41.9




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#20
Pentachlorophenol
Concen: 1.05 ng
RT: 16.81 min Scan# 1094
Delta R.T. -0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Tgt Ion:266 Resp: 531
Ion Ratio Lower Upper
266 100
268 77.5 62.5 93.7
264 57.3 57.2 85.8



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E





#21

Phenanthrene

Concen: 1.67 ng

RT: 17.18 min Scan# 1110

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :





Tot Ion:178 Resp: 14677

Ion Ratio Lower Upper

178 100

176 20.4 15.8 23.8

179 14.9 16.0 24.0#



#22

Anthracene

Concen: 1.46 ng

RT: 17.27 min Scan# 1114

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59




Tgt Ion:178 Resp: 12987


Ion Ratio Lower Upper

178 100

176 19.7 15.0 22.4

179 14.7 12.2 18.2





#23

Fluoranthene

Concen: 1.49 ng

RT: 19.18 min Scan# 1198

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :


Tot Ion:202 Resp: 63863

Ion Ratio Lower Upper

202 100

101 3.1 2.2 3.4




203 17.6 13.7 20.5



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#24

Chrysene-d12

Concen: 5.00 ng

RT: 21.30 min Scan# 1320

Delta R.T. -0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59


Tgt Ion:240 Resp: 42996

Ion Ratio Lower Upper

240 100

120 1.4 2.6 4.0#



236 21.2 24.6 37.0#




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#25
 Pvrene
 Concen: 1.79 ng
 RT: 19.54 min Scan# 1219
 Delta R.T. -0.00 min
 Lab File: BE092832.D
 Acq: 28 Mar 2017 11:59




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:202 Resp: 69780
 Ion Ratio Lower Upper
 202 100
 200 5.0 4.2 6.4
 203 17.2 13.9 20.9



#26
 Terphenyl-d14
 Concen: 2.25 ng
 RT: 19.76 min Scan# 1232
 Delta R.T. -0.00 min
 Lab File: BE092832.D
 Acq: 28 Mar 2017 11:59

Tgt Ion:244 Resp: 11665
 Ion Ratio Lower Upper
 244 100
 212 7.7 6.7 10.1
 122 2.3 3.6 5.4#









#27
Benzo(a)anthracene
Concen: 0.37 ng
RT: 21.26 min Scan# 1318
Delta R.T. -0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59




Instrument :
BNA_E
ClientSampled :


Tot Ion:228 Resp: 16017
Ion Ratio Lower Upper
228 100
226 32.4 23.6 35.4
229 16.6 13.3 19.9



#28
Chrysene
Concen: 0.44 ng
RT: 21.31 min Scan# 1321
Delta R.T. -0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Tgt Ion:228 Resp: 17206
Ion Ratio Lower Upper
228 100
226 35.4 36.3 54.5#
229 16.4 10.6 16.0#





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.88 ng

RT: 21.23 min Scan# 1316

Delta R.T. 0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59

Instrument :

BNA_E

ClientSampleId :


Tot Ion:149 Resp: 4145

Ion Ratio Lower Upper




149 100

167 26.1 16.0 24.0#

279 2.4 16.0 24.0#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#30

Indeno(1,2,3-cd)pyrene

Concen: 1.52 ng

RT: 26.25 min Scan# 1703

Delta R.T. 0.00 min

Lab File: BE092832.D

Acq: 28 Mar 2017 11:59


Tgt Ion:276 Resp: 65001

Ion Ratio Lower Upper



276 100


138 32.0 20.3 30.5#

277 24.8 19.7 29.5



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#31
Pervlene-d12
Concen: 5.00 ng
RT: 23.71 min Scan# 1492
Delta R.T. 0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59




Instrument :
BNA_E
ClientSampleId :


Tot Ion:264 Resp: 37502
Ion Ratio Lower Upper
264 100
260 25.4 20.1 30.1
265 22.4 17.9 26.9



#32
Benzo(b)fluoranthene
Concen: 1.60 ng
RT: 22.97 min Scan# 1428
Delta R.T. 0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Tgt Ion:252 Resp: 16029
Ion Ratio Lower Upper
252 100
253 24.1 18.7 28.1
125 12.4 10.5 15.7









#33
Benzo(k)fluoranthene
Concen: 1.43 ng
RT: 23.02 min Scan# 1432
Delta R.T. 0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59




Instrument :
BNA_E
ClientSampleId :


Tot Ion:252 Resp: 14355
Ion Ratio Lower Upper
252 100
253 22.2 20.3 30.5
125 12.6 9.6 14.4



#34
Benzo(a)pyrene
Concen: 1.44 ng
RT: 23.60 min Scan# 1482
Delta R.T. 0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Tgt Ion:252 Resp: 13095
Ion Ratio Lower Upper
252 100
253 21.8 19.0 28.6
125 15.1 11.8 17.8









#35
Dibenzo(a,h)anthracene
Concen: 1.63 ng
RT: 26.29 min Scan# 1705
Delta R.T. 0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Instrument :
BNA_E
ClientSampleId :

Tot Ion:278 Resp: 14172
Ion Ratio Lower Upper
278 100
139 18.0 13.8 20.8
279 24.5 21.4 32.2



#36
Benzo(g,h,i)perylene
Concen: 1.56 ng
RT: 27.02 min Scan# 1748
Delta R.T. 0.00 min
Lab File: BE092832.D
Acq: 28 Mar 2017 11:59

Tgt Ion:276 Resp: 57753
Ion Ratio Lower Upper
276 100
277 23.3 19.8 29.8
138 29.8 19.8 29.6#

